

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THQ HOSPITAL BUREWALA DISTRICT VEHARI**

**TECHNICAL EVALUATION FOR THE PURCHASE OF Monitor with Pulse, NIBP, ECG& etcO2
QTY-03, ESTIMATED COST RS. 100,000/-**

Sr# 1B- Monitor with Pulse, NIBP, ECG& etcO2

PART-I

Knock Down Criteria (Bid Evaluation)

Sr#	Evaluation Parameters	M/s Carevell	M/s Sial Traders	M/s Vital Care	M/s Radiant Medical	M/s Med Impex
1	Complete Package / Tender	Yes	Yes	Yes	Yes	Yes
2	Original Receipt of Tender	Yes	Yes	Yes	Yes	Yes
3	Affidavit from Bidder	Yes	Yes	Yes	Yes	Yes
4	Bid Security	Yes	Yes	Yes	Yes	Yes
5	Bid Validity	Yes	Yes	Yes	Yes	Yes
6	Delivery Period	Yes	Yes	Yes	Yes	Yes
	Remarks	Eligible	Eligible	Eligible	Eligible	Eligible

PART-II

Knock Down Criteria (Vendor Evaluation)

1	Valid Authorization Certificate from Manufacturer					
2	Technical & Engineering Capability	Yes	Yes	Yes	Yes	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes	Yes	Yes	Yes	Yes
4	Vendor Past Performance of Quoted Equipment					
5	Availability of relevent tools and testing / calibraion equipment	Yes	Yes	Yes	Yes	Yes
6	Compliance of Warranty as per tender	Yes	Yes	Yes	Yes	Yes
	Remarks					

PART III

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

Specification Compliance / Evaluation Parameters

Knock Down Criteria (Product Evaluation)

1	Model	XP Series-600.04-1	M7610	VIZOR-12S	Surveyor S-12	Datalys 780
2	Make	IGM&E GmbH	Oricare	Heyer	Welch Allyn	Lutech Medical
3	Country of Origin: USA, EEC, JAPAN	Germany	USA	Germany	USA	USA
4	Certification	CE	??	CE	??	CE
5	TECHNICAL SPECIFICATIONS:					
6	Bedside monitor for Adult, Paeds:	Yes	Yes	Yes	Yes	Yes
7	OPERATING FEATURES AND CHARACTERISTICS:					
8	Non-fade Touch screen TFT color display	TFT But Touch screen?	TFT Touch	TFT Touch	TFT Touch	TFT Touch
9	Electro-surgical interference suppression / protection.	Yes	Yes	Yes	??	Yes
10	Defibrillator protection.	Yes	Yes	Yes	??	Yes
11	Freeze and cascade facility.	??	??	Yes	??	Yes
12	Waveform traces speed: 25 / 50mm/sec.	50mm/sec ?	??	Yes	Yes	Yes
13	Screen size: min. 12" TFT / LCD Touch Screen color display.	12" / Touch?	15" TFT	12" TFT	12" TFT	12.1" TFT
14	Capacity to interface with LAN/WLAN for data transfer.	Yes	Yes	Yes	Yes	Yes
15	ECG:					
16	Numeric: heart rate.	Yes	Yes	Yes		Yes
17	Waveform: Six Wave forms minimum, real time and freeze ECG trace.	13 waveform but no freeze?	13 waveform but no freeze?	08 Waveform	08 waveform but no freeze?	12 Waveforms
18	NON-INVASIVE BLOOD PRESSURE(NIBP):					
19	Method:Oscillometric principle	Oscillometric	Oscillometric	Oscillometric	Oscillometric	Oscillometric
20	Numeric: systolic, diastolic and mean pressure	Yes	Yes	Yes	Yes	Yes
21	Selectable auto inflates interval settings.	Yes	Yes	Yes	Yes	Yes
22	Rising cuff continuous pressure display.	Yes	Yes	Yes	Yes	Yes
23	Reusable cuff	Offered	Offered	Offered	Offered	Offered

[Handwritten signatures]

[Handwritten signatures]

[Handwritten signature]

24	TEMPERATURE:							
25	Numeric: temperature selectable in °C/°F.	Yes	Yes					Yes
26	PULSE OXIMETRY:							
27	Numeric: 0-100% oxygen saturation measuring range.	Yes	Yes					Yes
28	Waveform-Plethysmograph pulse.	Yes	Yes					Yes
29	Reusable sensor electrode.	Offered	Offered					Offered
30	ARRHYTHMIA ANALYSIS:							
31	Arrhythmia analysis and ST analysis	Yes	Yes					Yes
32	RESPIRATION:							
33	Breath rate display and apnea alarms.	Yes	Yes					Yes
34	Sweep speed: 6.25, 12.5 mm/sec.	Yes	Yes					Yes
35	OTHER FEATURES:							
36	Trend data: graphical and tabular.	Yes	Yes					Yes
37	ALARMS:							
38	High & low (settable) on all parameters.	Yes	Yes					Yes
39	Visual and audible indication of alarms.	Yes	Yes					Yes
40	OPERATING REQUIREMENTS:							
41	AC 220V/50HZ	Yes	Yes					Yes
42	Built-in rechargeable battery for at least 1 - 2hour AC power failure at full parameter.	??	4-Hours		120 Minutes		3-Hours	4-Hours
43	NOTE: The system must be complete with all sensors, probes, cables or any other accessories required for measuring all the above selected parameters for Adults and Peds.	Offered	Offered		Offered		Offered	Offered
44	ACCESSORIES:							
45	Wall Mounting stand and Mobile with Basket preferably imported or high quality (S.S 304L) Local with lockable draws (procuring agency to choose)	Optional	Not Offered		Offered		Offered	Offered
46	OPTIONAL:							

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

47	IBP Two Channel / Four Channel.	4-Channel	2-Channel	2-Channel	2-Channel	2-Channel
48	Capnography (ETCO2).	Offered	Offered	Offered	Offered	Offered
49	Cardiac output.	Offered	Offered	Offered	Offered	Offered
50	Single / Dual Channel Printer	3-Channel	??	2-Channel	2-Channel	2-Channel
	SPECIFICATION COMPLIANCE	No	No	Yes	No	Yes
	TECHNICAL ELIGIBILITY OF THE PRODUCT	No	No	Yes	No	Yes
	TECHNICAL ELIGIBILITY OF THE FIRM	Yes	Yes	Yes	Yes	Yes
	Bid Status	Non-Responsive	Non-Responsive	Responsive	Non-Responsive	Responsive

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THQ HOSPITAL BUREWALA DISTRICT VEHARI**

**TECHNICAL EVALUATION FOR THE PURCHASE OF CENTRAL OXYGEN SUPPLY
QTY-01, ESTIMATED COST RS. 100,00,000/-**

S# 2B - CENTRAL OXYGEN SUPPLY

PART-I

Knock Down Criteria (Bid Evaluation)

Sr#	Evaluation Parameters	M/s Total Technologies	M/s Eastern Medical	M/s Intexim	M/s Radiant Medical	M/s Carewell
1	Complete Package / tender	No	Yes	No	Yes	Yes
2	Original Receipt of Tender	Yes	Yes	Yes	Yes	Yes
3	Affidavit from Bidder	Yes	Yes	Yes	Yes	Yes
4	Bid Security	Yes	Yes	Yes	Yes	Yes
5	Bid Validity	Yes	Yes	Yes	Yes	Yes
6	Delivery Period	Yes	Yes	Yes	Yes	Yes
	Remarks	Not-Eligible	Eligible	Not-Eligible	Eligible	Eligible

PART-II

Knock Down Criteria (Vendor Evaluation)

1	Valid Authorization Certificate from Manufacturer	Yes	Yes	Yes	Yes	Yes
2	Technical & Engineering Capability	Yes	Yes	Yes	Yes	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes	Yes	Yes	Yes	Yes
4	Vendor Past Performance of Quoted Equipment	Yes	Yes	Yes	Yes	Yes
5	Availability of relevant tools and testing / calibration equipment	Yes	Yes	Yes	Yes	Yes
6	Compliance of Warranty as per tender	Yes	Yes	Yes	Yes	Yes
	Remarks	Eligible	Eligible	Eligible	Eligible	Eligible

PART III

Knock Down Criteria (Product Evaluation)






Specification Compliance / Evaluation Parameters

1	Model	MCU 100	10366H-2002M	MQ 180D	Pure Gas Part No. 8102341320	No Data Sheet Provided
2	Make	Medgas Technik	Deltap	Medicop	Pneumatic Medical Gas Solution By Atlas Copco	
3	Country of Origin	Germany	Italy	Slovenia	UK/ Europe	
4	Certification	CE	CE	CE	CE	
5	Technical Specification	The Manifold shall be a fully automatic type & shall switch from "Bank in use" to "Reserve bank" without fluctuation in delivery of supply line pressure & without the need for external power.				
6		Yes	Yes	Yes	Yes	
7	After switch over the "Reserve bank" it will become the "Bank in use" and other will become the "Reserve bank". The system should be able to maintain continuous supply in case of power failure.	Yes	Yes	Yes	Yes	
8	AUTOMATIC-MANIFOLD FOR OXYGEN	MCU 100	10366H-2002M	MQ 180D	Part No. 8102341320	
9	CAPACITY: (2 x 8) with capacity of 100-120 m3/h.	(2 x 8) with capacity of 120 m3/h.	(2 x 8) with capacity of 120 m3/h.	(2 x 10) with capacity of 180 m3/h.	(2 x 6) with capacity of 105 m3/h.	
10	FUNCTION:					
11	Fully automatic self-contained shuttle-valve with no electrical power required for switching.	Fully automatic	Fully automatic	Fully automatic	Fully automatic	
12	Microprocessor controlled display unit for monitoring of different parameters and loggings.	Microprocessor Controlled	Microprocessor Controlled	Microprocessor Controlled	Microprocessor Controlled	
13	Units of pressure measurement (Psi / Pa/ Bar).	All	All	All	All	
14	LCD /LED display for the left bank, the right bank & for the supply pressure.	LCD Display	LED Display	LCD Display	LCD Display	

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

15	Complete with relief valve, Emergency shut off valve and battery / bank room alarm for left and right cylinder bank.	Yes	Yes	Yes	Yes	Yes	
16	VIE Connection with shut off valve, controlled by the manifold cabinet.	Yes	Yes	Yes	Yes	Yes	
17	The primary source of Oxygen will be the VIE System.	VIE Connection	VIE Connection	Part No. 1070023	VIE Connection		
18	The complete automatic manifold unit with shifting from VIE System to Main Bank and then to Reserve Bank automatically with display of pressures for each source.	Yes	Yes	Yes	Yes	Yes	
19	Equipped with NIST(Non-Interchangeable Screw Threaded) Connection.	Yes	Yes	Yes	Yes	Yes	
20	AUTOMATIC-MANIFOLD FOR NITROUS OXIDE						
	Model	Not Quoted	11091H-2002M	Not Quoted	Pure Gas Part No. 8102341321		
	Make	Not Quoted	Deltap	Not Quoted	Pneumatic Medical Gas Solution By Atlas Copco		
	Country of Origin	Not Quoted	Italy	Not Quoted	UK/ Europe		
21	CAPACITY: (2 x 4), with capacity of 70-90, m3/h	Not Quoted	(2 x 4) with capacity of 120 m3/h.	Not Quoted	(2 x 4) with capacity of 105 m3/h.		
22	FUNCTION:						
23	Fully automatic self-contained shuttle-valve with no electrical power required for switching.	Not Quoted	Yes	Not Quoted	Yes		
24	Microprocessor controlled display unit for monitoring of different parameters and loggings.	Not Quoted	Yes	Not Quoted	Yes		
25	Units of pressure measurements (Psi / Pa / Bar).	Not Quoted	Yes	Not Quoted	Yes		
26	LCD / LED display for the left bank, the right bank for the supply pressure.	Not Quoted	LED Display	Not Quoted	LED Display		

27	Complete with relief valve, Emergency shut off valve and battery / bank room alarm for left and right cylinder bank equipped with NIST Connection	Not Quoted	Yes	Not Quoted	Yes	
28	BED HEADS UNIT FOR ICU (Qty-??)					
29	Model	Not Quoted	EVA XL	Not Quoted	Customized	
30	Make	Not Quoted	Deltap	Not Quoted	Pneumatic Medical Gas Solution By Atlas Copco	
31	Country of Origin	Not Quoted	Italy	Not Quoted	UK/ Europe	
32	Each unit consists of: 1,500 mm or above.	Not Quoted	1800 mm	Not Quoted	1800 mm	
33	Horizontal type with railing.	Not Quoted	Horizontal	Not Quoted	Horizontal	
34	Built in over bed Light with ON/OFF switch.	Not Quoted	Yes	Not Quoted	Yes	
35	Built in reading Light with ON/OFF switch.	Not Quoted	Yes	Not Quoted	Yes	
36	Provision for Nurse Call/ Alert Button	Not Quoted	Nurse Call	Not Quoted	Nurse Call	
37	Gas outlets (1x Vacuum, 2x Air, 2x Oxygen complete fitted).	Not Quoted	1x Vacuum, 2x Air, 2x Oxygen complete fitted	Not Quoted	1x Vacuum, 2x Air, 2x Oxygen complete fitted	
38	Electrical sockets: 6 Nos. & 2Nos. Multi Pin Plug and above schuko / F-type	Not Quoted	6 Nos. & 2Nos. Multi Pin Plug and above schuko	Not Quoted	6 Nos. & 2Nos. Multi Pin Plug and above schuko	
39	All wiring conforms to international standards,	Not Quoted	Yes	Not Quoted	Yes	
40	Groundearth	Not Quoted	Yes	Not Quoted	Yes	
41	Monitor shelf with railing clamp	Not Quoted	Monitoring Shelf	Not Quoted	Monitoring Shelf	
42	IV pole with railing clamp	Not Quoted	??	Not Quoted	??	
43	Separate ducts for Electrical & Gas Pipes.	Not Quoted	Yes	Not Quoted	Yes	
44	BED HEADS UNIT FOR WARDS & PRIVATE ROOM (Qty-??)					

[Handwritten signatures and initials]

[Handwritten signature]

[Handwritten signature]

45	Model	Care Series (Care BLAKOU 1000)	EVA XL	Medicompact Double Type	Customized
46	Make	Medgas Technik	Deltap	Medicop	Pneumatic Medical Gas Solution By Atlas Copco
47	Country of Origin	Germany	Italy	Slovenia	UK/ Europe
48	Each unit consists of: 1,000 mm and above	1000 mm	1800 mm	1000 mm	1800 mm
49	Horizontal type	Horizontal Type	Horizontal Type	Horizontal Type	Horizontal Type
50	Built in over bed Light with ON/OFF switch.	Yes	Yes	Yes	Yes
51	Built in reading Light, with ON/OFF switch.	Yes	Yes	Yes	Yes
52	Provision for Nurse Call Alert Button	Yes	Yes	Yes	Yes
53	Gas outlets (1x Vacuum, 1x Air, 2x Oxygen complete fitted).	1x Vacuum, 1x Air, 2x Oxygen	1x Vacuum, 1x Air, 2x Oxygen	1x Vacuum, 1x Air, 2x Oxygen	1x Vacuum, 1x Air, 2x Oxygen
54	Electrical sockets: 4 Nos. & 2Nos. Multi Pin Plug and above schuko/ F-type.	4 Nos. & 2Nos. Multi Pin Plug	4 Nos. & 2Nos. Multi Pin Plug	4 Nos. & 2Nos. Multi Pin Plug	4 Nos. & 2Nos. Multi Pin Plug
55	All wiring conforms to international standards.	Yes	Yes	Yes	Yes
56	Groundearth.	Yes	Yes	Yes	Yes
57	Separate ducts for Electrical & Gas Pipes.	Yes	Yes	Yes	Yes
58	ALARM AREA (Qty-??)				
59	Model	TMA 0701			Zeus 6
60	Make	Telemedica			Pneumatic Medical Gas Solution By Atlas Copco
61	Country of Origin	Italy	Italy	Slovenia	UK/ Europe

Multi Zone Valve Box

Multi Zone Valve Box

Disanti

Amor

L-T

Amor

62	Alarm area for two and three gases	3-Gases not Quoted				Offered Both	
63	Individual display and sensor module.	Yes				Yes	
64	Self-diagnostic & error message display readable for ease of maintenance.	Yes				Yes	
65	Psl, Pa or Bar read out	Yes				Yes	
66	Test & Mute Button.	Yes				Yes	
67	Low alarm condition.	Yes				Yes	
68	MULTIPLE ZONE VALVE BOX (Qty-??)						
69	Model	ZVB	11041/11042 (2620D)	Cat No. 1021221	??		
70	Make	Bacon Medas	Deltap	Medicop	Pneumatic Medical Gas Solution By Atlas Copco		
71	Country of Origin	UK	Italy	Slovenia	UK/ Europe		
72	Multiple zone valve box for two and three gases	3 Gases not Quoted	2- Gases (11041-2620D) 3- Gases (11042-2620D)	3-Gases not Quoted	Offered Both		
73	Each wall type zone valve box shall consist of the following Components:	Yes	Yes	Yes	Yes		
74	A steel valve box which can house two to six shutoff valves.	Yes	Yes	Yes	Yes		
75	Pressure gauges to display the pressure of various gases.	Yes	Yes	Yes	Yes		
76	OXYGEN FLOW METER (Qty-??)						
77	Model	RS+CH 200	8532B-2020	RS	Part No. 5540002		
78	Make	Flow Meter	Deltap	Flow Meter	Pneumatic Medical Gas Solution By Atlas Copco		

Handwritten signature and notes in blue ink.

Handwritten signature in blue ink.

Handwritten signature in blue ink.

79	Country of Origin	Italy	Italy	Italy	UK/ Europe
	Valid Authorization Certificate from Manufacturer	Provided	Provided	Not Provided	Yes
80	Oxygen flow meter with humidifier and probes.	Yes	Yes	Yes	Yes
81	Oxygen Flow meter complete set from 1 to 15 LPM	Yes	Yes	Yes	Yes
82	With autoclave able and unbreakable humidifier bottle.	Yes	Yes	Yes	Yes
83	For neonate / Paeds (0-6 LPM)	Yes	??	Yes	Yes
84	SUCTION VENTURI				
85	Model	Easy Air with Mac 2000	Not Quoted	AV 500	Part No. 4430021
86	Make	Flow Meter	Not Quoted	Flow Meter	Pneumatic Medical Gas Solution By Atlas Copco
87	Country of Origin	Italy	Not Quoted	Italy	UK/ Europe
88	Sterilizable Bottle	Not Offered ??	Not Quoted	MAK-1000	Yes
89	COPPER PIPING				
90	Make	Tonoli	Dellap	Imported	Lawton
91	Country of Origin	Italy	Italy	??	UK
92	Supply and installation of seamless medical graded copper pipe. Deoxidized and degreased along with required fitting etc.	Yes	Yes	Yes	Yes
93	Various sizes / diameter as per Drawing of the project and Design standard of the bidder with matching color indications of our lists.	Yes	Yes	Yes	Yes
94	The sizes will vary from minimum of 10mm to onward as per design offered by the firm	Yes	Yes	Yes	Yes
95	COLOR CODING	Yes	Yes	Yes	Yes
96	Oxygen: White	Yes	Yes	Yes	Yes
97	Air: Black	Yes	Yes	Yes	Yes
98	Nitrous: Blue	Yes	Yes	Yes	Yes
99	Vacuum: Yellow	Yes	Yes	Yes	Yes



W. Sharma - /

[Signature]

[Signature]

MEDICAL COMPRESSED AIR STATION:						
100	MEDICAL COMPRESSED AIR STATION:					
101	Medical Air Plant with Breathable Air. Elimination of Toxic Gases such as carbon dioxide, Nitrogen dioxide, Nitrogen monoxide, Carbon monoxide; and other pollutants as per European	Yes	Yes	Yes	Yes	
102	COMPRESSOR					
103	Model	SX8	CA0068-2047	D-VL-13 845	MedAir2455-10GD	
104	Make	Kaesser	Dellap	Novair Medical	Pneumatic Medical Gas Solution By Atlas Copco	
105	Country of Origin	Germany	Italy	France	UK/ Europe	
106	Duplex Compressor unit	Duplex unit	Duplex unit	Duplex unit	Duplex unit	
107	Reciprocating / Screw / Scroll type.	Screw	Screw	Screw	Screw	
108	Capacity of air producing approximately minimum of each compressor - 2500 liters/min / FAD @ 13 bars.	0.54 m3/h FAD@13 Bars	15 - 400 Liters/Min FAD@13 Bars	36 m3/h FAD@13 Bars	2850 Liters/Min FAD@13 Bars	
109	Mounted on anti-vibration base.	Yes	Yes	Yes	Yes	
110	RESERVOIR					
111	Compatible medical grade.	Yes	Yes	Yes	Yes	
112	Air receiving tank 1,000 L. or as per actual requirement.	300 Liters	1000 Liters	1000 Liters	1500 Liters	
113	Vertical type.	Yes	Yes	Yes	Yes	
114	Galvanized internally with auto drain.	??	??	??	??	
115	Minimum 13 bars out put pressure of tank.	Yes		Yes	Yes	
116	AIR DRYER					
117	Model	MT 6	??	DATS-13	Pure Med 35	

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

118	Make						
119	Country of Origin	Germany	Italy	France	UK/ Europe		
120	Desiccant type, Duplex.	Duplex unit	??	Duplex	Duplex		
121	Dew Point range between -40 to -80°C	??	??	-70 Degree Centigrade	-46 Degree Centigrade		
122	Dew point monitoring on LCD Panel	??	??	LCD Panel	LCD Panel		
123	Capacity suitable according to the compressor output.	??	??	??	??		
124	Including oil water separator.	??	??	Included	Included		
125	FILTRATION SYSTEM						
126	Clean Medical Grade Air supply in accordance with the requirements of HTM/ISO standards.	Yes	Yes	Yes	Yes		
127	Consisting of Pre-filter / humidity	Yes	Yes	Yes	Yes		
128	Oil free and sterile/bacteria filter.	Yes	Yes	Yes	Yes		
129	Mounted with shut-off valves on an assembly panel.	Yes	Yes	Yes	Yes		
130	Parallel Connections of the filters. This will make it possible to exchange filter without interrupting the air supply.	Yes	Yes	Yes	Yes		
131	Flow / filtering rate according to the compressed air output.	Yes	Yes	Yes	Yes		
132	REDUCER PANEL COMPRESSED AIR						
133	Parallel switched reducer with gauge, safety valve, pressure switch for high and low Pressure and shut off valve with assembly panel size 4 + 7 bar.	Yes	Yes	Yes	Yes		
134	Complete with distribution block according to requirement, including Shut off valve and pressure gauge for every distribution block.	Yes	Yes	Yes	Yes		
135	Complete assemble panel with incoming and outgoing copper pipe for complete system.	Yes	Yes	Yes	Yes		
136	CONTROL						

Handwritten signatures and initials:
 [Signature] [Signature] [Signature]

Handwritten signature:
 [Signature]

Handwritten signature:
 [Signature]

137	Complete with 1x main warning system for compressed air for visual and acoustic monitoring of alarms conditions for the compressor room.	Yes	Yes	Yes	Yes	Yes	Yes	
138	Test point in the system for air quality.	Yes	Yes	Yes	Yes	Yes	Yes	
139	1x switch cabinet for automatic Unit including.	Yes	Yes	Yes	Yes	Yes	Yes	
140	All necessary fitting and installation material for smooth running of the system without any interruption.	Yes	Yes	Yes	Yes	Yes	Yes	
141	POWER:							
142	Power Supply- 3 Phase, 390 - 400V/50 Hz.	Yes	Yes	Yes	Yes	Yes	Yes	
143	SPECIAL TERMS & CONDITIONS:							
144	The Major equipment like manifold, pendants, bed head units and outlets should be manufactured and supplied by the same Principal.	??	Same Manufacturer	??	Same Manufacturer	??	Same Manufacturer	
145	The Vacuum System and medical air compressor may be of different manufacturer but should be supplied by the same manufacturer of medical gas pipeline system.	Offered	Offered	Offered	Offered	Offered	Offered	
146	Separate ducts for piping and electrification.	Yes	Yes	Yes	Yes	Yes	Yes	
147	All equipment must be according to international safety standard.	Yes	Yes	Yes	Yes	Yes	Yes	
148	The color of outlets, piping and cylinder for recognition will be BS type for all the system equipment.	Yes	Yes	Yes	Yes	Yes	Yes	
149	The drop outlet/ Connection will be flushed surface.	Yes	Yes	Yes	Yes	Yes	Yes	
150	Country of Manufacturer: USA, EU AND JAPAN	Germany	Italy	France	UK/EEC			
	SPECIFICATION COMPLIANCE	No	No	No	No	No	No	No
	TECHNICAL ELIGIBILITY OF THE PRODUCT	No	No	No	No	No	No	No
	TECHNICAL ELIGIBILITY OF THE FIRM	Yes	Yes	Yes	Yes	Yes	Yes	Yes
	Bid Status	Non-Responsive	Non-Responsive	Non-Responsive	Non-Responsive	Non-Responsive	Non-Responsive	Non-Responsive

Handwritten signature

Handwritten signature

Handwritten signature

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THQ HOSPITAL BUREWALA DISTRICT VEHARI**

**TECHNICAL EVALUATION FOR THE PURCHASE OF Anesthesia Machine with Digital Flow Meter & Integrated Ventilator
QTY-02, ESTIMATED COST Rs. 22,00,000/- EACH**

Sr# 3B - Anesthesia Machine with Digital Flowmeter & Integrated Ventilator

PART-I

Knock Down Criteria (Bid Evaluation)

Sr#	Evaluation Parameters	M/s Eastern Medical	M/s Radiant Medical	M/s Sial Traders
1	Complete Package / tender	Yes	Yes	No
2	Original Receipt of Tender	Yes	Yes	Yes
3	Affidavit from Bidder	Yes	Yes	Yes
4	Bid Security	Yes	Yes	Yes
5	Bid Validity	Yes	Yes	Yes
6	Delivery Period	Yes	Yes	Yes
Remarks		Eligible	Eligible	Not-Eligible

PART-II

Knock Down Criteria (Vendor Evaluation)

1	Valid Authorization Certificate from Manufacturer	Yes	Yes	Yes
2	Technical & Engineering Capability	Yes	Yes	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes	No	No
4	Vendor Past Performance of Quoted Equipment	Yes	Yes	Yes
5	Availability of relevant tools and testing / calibration equipment	Yes	Yes	Yes
6	Compliance of Warranty as per tender	Yes	Yes	Yes
Remarks		Eligible	Not-Eligible	Not-Eligible

PART III

Knock Down Criteria (Product Evaluation)

Specification Compliance / Evaluation Parameters

	Specification Compliance / Evaluation Parameters	Product	Model	Make
1		Pasithec	AX 700	A9600
2		Hoyer Medical	Dameca	Oricare

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

3	Country of Origin	Germany	Denmark	USA
4	Certification	CE	CE	CE
5	Technical Specification			
6	Anesthesia machine to administer anesthetic agents in precise control and flow manner for adult, pediatric and neonates.	Neonates??	Neonates??	Neonates??
7	The machine will equip to monitor the vital sign parameters and anesthetic agents during operation.	Vital Sign with agent monitoring	Vital Sign with agent monitoring	Vital Sign with agent monitoring
8	Mobile 3-gases O2 /N2O /AIR.	Mobile 3-Gases	??	Mobile 3-Gases
9	It must comprise of the following components:			
10	Non-interchangeable pipeline inlets.	NIST Connection	NIST Connection	NIST Connection
11	Pipeline & cylinder gauges for O2, N2O and AIR.	Yes	Yes	Yes
12	Central gas / electronically driven unit.	Pneumatically Driven	Pneumatically Driven	Pneumatically Driven
13	Pin index cylinder yokes for Oxygen & N2O (One each), as backup.	3-Cylinders	??	??
14	Pin index type Cylinders will be provided (2xO2 and 2xN2O: BS standard).	Offered	Offered	Offered
15	Gas outlet and O2 flush control.	Yes	Yes	Yes
16	1 auxiliary O2 outlet.	Yes	Yes	Yes
17	Two Lockable castors	2-Lockable Castors	2-Lockable Castors	2-Lockable Castors
18	Stainless steel/fiber work surface	Fiber Work Surface	Fiber Work Surface	Fiber Work Surface
19	Absorber bag support arm	GCX Rail	Yes	GCX Rail
20	Three gas digital flow meters for precise control and monitoring of gases and virtual tube display.	Virtual tube display	Digital flowmeter / Virtual Tube??	Virtual tube display
21	Drawer unit 4- 6" high	??	??	??
22	Scavenging system Passive / Active type	??	AGSS	??
23	ANESTHESIA VENTILATOR:			
24	Anesthesia Ventilator with minimum 10" or more color LCD/TFT screen.	12.1" Color Touch	15" Color Touch	12.1" Color Touch
25	The ventilator shall be capable of ventilating adult and pediatric patients.	Yes	Yes	Yes
26	The ventilator shall have following features as a minimum requirement:	Yes	Yes	Yes
27	Volume Preset Time Cycled Ventilator (IPV Model)	Yes	??	Yes
28	Pressure Controlled and pressure support Modes	Yes	Yes	Yes
29	Breathing Mode Selection (Standby / Volume / Spontaneous and Pressure)	Yes	Yes	Yes

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

30	Built in Oxygen Monitor	Yes	Yes	Yes
31	Inverse I:E ratio Capability	Yes	Yes	Yes
32	Gas Specific Input Connectors (Air or Oxygen ISO or ANSI Standards)	Yes	Yes	Yes
33	Tidal Volume from (5ml to 1400 ml)	20 --1500ml	05--1500ml	20 --1500ml
34	Rate or Frequency 4 to 60 bpm	4--100bpm	4--80bpm	4--100bpm
35	PEEP (4 to 20 cm H2O)	3--30cmH2O	4--20cmH2O	3--30cmH2O
36	Inspiratory Pressure Limit	Yes	Yes	Yes
37	Power Supply 220 VAC, 50 Hz	Yes	Yes	Yes
38	Battery Backup (60 Minutes or more)	??	??	??
39	Low / High FIO2 Alarm	Yes	Yes	Yes
40	Incorrect Rate or Ratio alarm	Yes	Yes	Yes
41	Mains Failure alarm / Low battery alarm	Yes	Yes	Yes
42	Hypoxic device guard.	Yes	Yes	Yes
43	Pressure and Volume (Spirometry) Loops / curves.	Spirometry loop	Spirometry loop ??	Spirometry loop
44	High / Low pressure alarm.	Yes	Yes	Yes
45	The ventilator shall be supplied with complete drive hose and power cable.	Yes	Yes	Yes
46	Note: Annual maintenance kits (needs to replace annually) will be included in the warranty period as per manufacturer's guidelines.	Offered	Offered	Offered
47	VITAL SIGN MONITOR			
48	Model	VizOR 17	Expzezon	Not-Offered
49	Make	Heyer Medical	Spacelab Healthcare	
50	Country of Origin	Germany	USA	
51	TECHNICAL SPECIFICATION			
52	Size of minimum 15" Touch Screen or more for display of vital sign parameters.	18.5 TFT	19" Color Touch	
53	Measurement of ECG 5 leads.	Yes	Yes	
54	NIBP with re-usable single hose cuff for children and adults.	Yes	Yes	
55	SpO2 with re-usable cable and sensors for children and adults' size (Massimo Type / Equivalent motion tolerance technology).	Yes	Yes	
56	HR.	Yes	Yes	
57	Temperature with nasal probe.	Yes	Yes	
58	Respiration.	Yes	Yes	







59	Dual Channel IBP.	Yes	Yes	
60	220V, 50 Hz operated.	Yes	Yes	
61	Note: Vial sign Monitor must be supplied by the same manufacturer and must be Compatible with the machine and Ventilator.	Same Manufacturer	Specelab Healthcare	
62	The warranty of equipment will be including batteries, oxygen sensor and flow sensor.	Offered	Offered	
63	Anesthesia Accessories:			
64	Power outlet with 3/4 socket outlets to connect the auxiliary equipment	Offered	Offered	
65	CO2 absorber 800-1,500 gm or better complete with valve for bag/ventilator.	Offered	Offered	
66	Manometer.	Offered	Offered	
67	Breathing bags.	Offered	Offered	
68	Re-usable Silicon Autoclavable breathing circuit (Adult, Peds, Infant 01 each).	Offered	Offered	
69	Mounts and Y-piece.	Offered	Offered	
70	Additional breathing hose and connector with 03adult& 03pediatricbellows.	Offered	Offered	
71	Monitor Accessories:			
72	1. 2 NIBP Cuff each	Offered	Offered	
73	2. 2 Spo2 probe.	Offered	Offered	
74	3. 2 nasal temperature probes.	Offered	Offered	
75	4. 2 ECG Leads.	Offered	Offered	
76	5. IBP Leads.	Offered	Offered	
77	Country of Manufacturer: USA, EU AND JAPAN	EU	EU/USA	USA
		SPECIFICATION COMPLIANCE		
		No	No	No
	TECHNICAL ELIGIBILITY OF THE PRODUCT	No	No	No
	TECHNICAL ELIGIBILITY OF THE FIRM	Yes	Yes	Yes
	Bid Status	non-Responsive	non-Responsive	non-Responsive

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THQ HOSPITAL BUREWALA DISTRICT VEHARI**

**TECHNICAL EVALUATION FOR THE PURCHASE OF Bed Side Cabinet
(Qty-90) Estimated Cost. Rs. 12,500/- Each**

Sr# 4B Bed Side Cabinet

PART-I

Knock Down Criteria (Bid Evaluation)

Sr#	Evaluation Parameters	M/s Pieces Engineering & GOS	M/s Thankful Agencies	M/s Noorani Surgical (Pvt) Ltd
1	Complete Package / tender	Yes	Yes	Yes
2	Original Receipt of Tender	Yes	Yes	Yes
3	Affidavit from Bidder	Yes	Yes	Yes
4	Bid Security	No	Yes	Yes
5	Bid Validity	Yes	Yes	Yes
6	Delivery Period	Yes	Yes	Yes
Remarks				
		Not-Eligible	Eligible	Eligible

PART-II

Knock Down Criteria (Vendor Evaluation)

1	Valid Authorization Certificate from Manufacturer	Yes	Yes	Yes
2	Technical & Engineering Capability	Yes	Yes	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes	Yes	Yes
4	Vendor Past Performance of Quoted Equipment	Yes	Yes	Yes
5	Availability of relevant tools and testing / calibration equipment	Yes	Yes	Yes
6	Compliance of Warranty as per tender	Yes	Yes	Yes
Remarks				
		Eligible	Eligible	Eligible

PART III

Knock Down Criteria (Product Evaluation)





Specification Compliance / Evaluation Parameters

1	Model	Local	Local	Local
2	Make	Local	Local	Local
3	Country of Origin	Pakistan	Pakistan	Pakistan
4	Technical Specification			
5	Powder coated sheet steel construction	Yes	Yes	Yes
6	Galvanized steel sheet 20 SWG	Yes	Yes	Yes
7	Two drawers with one bottom compartment	Yes	Yes	Yes
8	Built-in handles for durability	Yes	Yes	Yes
9	Plastic top laminated with raised edges on 3 sides	Yes	Yes	Yes
10	75mm rustproof, waterproof twin castors with two lockable castors	Yes	Yes	Yes
11	Castors with rubber tires for noise free operation on floors	Yes	Yes	Yes
12	Buffers on all four corners	Yes	Yes	Yes
13	Compatible with the offered bed in colour and functionality	Yes	Yes	Yes
14	All metal parts should be welded by CO2 gas	Yes	Yes	Yes
15	Size: 18"x15"x34" (WxDxH) approx.	Yes	Yes	Yes
16	Local Made Subject to approval of sample			
SPECIFICATION COMPLIANCE		Yes	Yes	N/A
TECHNICAL ELIGIBILITY OF THE PRODUCT		Yes	Yes	N/A
TECHNICAL ELIGIBILITY OF THE FIRM		NO	Yes	Yes
Bid Status		Non-Responsive	Responsive	Non-Responsive

[Handwritten signatures]

[Handwritten signatures]

Sample not provided

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THQ HOSPITAL BUREWALA DISTRICT VEHARI
TECHNICAL EVALUATION FOR THE PURCHASE OF Over Bed Table
(Qty-90) Estimated Cost. Rs. 7,000/- Each**

Srf# 05B Over Bed Table

PART-I**Knock Down Criteria (Bid Evaluation)**

Sr#	Evaluation Parameters	M/s Pieces Engineering & GOS	M/s Noorani Surgical (Pvt) Ltd	M/s Thankful Agencies
1	Complete Package / tender	Yes	Yes	Yes
2	Original Receipt of Tender	Yes	Yes	Yes
3	Affidavit from Bidder	Yes	Yes	Yes
4	Bid Security	No	Yes	Yes
5	Bid Validity	Yes	Yes	Yes
6	Delivery Period	Yes	Yes	Yes
	Remarks	Not-Eligible	Eligible	Eligible

PART-II**Knock Down Criteria (Vendor Evaluation)**

1	Valid Authorization Certificate from Manufacturer	Yes	Yes	Yes
2	Technical & Engineering Capability	Yes	Yes	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes	Yes	Yes
4	Vendor Past Performance of Quoted Equipment	Yes	Yes	Yes
5	Availability of relevant tools and testing / calibration equipment	Yes	Yes	Yes
6	Compliance of Warranty as per tender	Yes	Yes	Yes
	Remarks	Eligible	Eligible	Eligible

PART III**Knock Down Criteria (Product Evaluation)**





Specification Compliance / Evaluation Parameters

1	Model	Local	Local	Local
2	Make	Local	Local	Local
3	Country of Origin	Pakistan	Pakistan	Pakistan
4	Technical Specification			
5	Powder coated mild steel construction	Yes	Yes	Yes
6	Stainless steel internal column	Yes	Yes	Yes
7	Gas spring assisted height	Yes	Yes	Yes
8	Height adjustable from 70 to 100cm adjustments	Yes	Yes	Yes
9	Plastic top laminated with rounded edges and cup/glass holder	Yes	Yes	Yes
10	Approx top size: 80x32cm	Yes	Yes	Yes
11	50mm twin castors	Yes	Yes	Yes
12	Castors with rubber tires for noise free operation on floors	Yes	Yes	Yes
13	Compatible with the offered bed in colour and functionality	Yes	Yes	Yes
14	All metal parts should be welded by CO2 gas	Yes	Yes	Yes
15	Base covered with stainless steel sheet for use as foot rest	Yes	Yes	Yes
16	Local Made Subject to approval of sample			
SPECIFICATION COMPLIANCE		Yes	No	Yes
TECHNICAL ELIGIBILITY OF THE PRODUCT		Yes	No	Yes
TECHNICAL ELIGIBILITY OF THE FIRM		No	Yes	Yes
Bid Status		Non-Responsive	Non-Responsive	Responsive

[Handwritten signatures and notes]

[Handwritten signature]
[Handwritten signature]
 Mr. Sultan 2

[Handwritten signature]

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THQ HOSPITAL BUREWALA DISTRICT VEHARI
TECHNICAL EVALUATION FOR THE PURCHASE OF ELECTRIC DIATHERMY 400 WATT
QTY-02, ESTIMATED COST RS. 800,000/- EACH**

SIT# 68 - ELECTRIC DIATHERMY 400 WATT

PART-I

Knock Down Criteria (Bid Evaluation)

S#	Evaluation Parameters	M/s G-Med	M/s Vital Care	M/s Sigma International	M/s Mediland
1	Complete Package / tender	Yes	Yes	Yes	Yes
2	Original Receipt of Tender	Yes	Yes	Yes	Yes
3	Affidavit from Bidder	Yes	Yes	Yes	Yes
4	Bid Security	Yes	Yes	Yes	Yes
5	Bid Validity	Yes	Yes	Yes	Yes
6	Delivery Period	Yes	Yes	Yes	Yes
Remarks					
Eligible					

PART-II

Knock Down Criteria (Vendor Evaluation)

1	Valid Authorization Certificate from Manufacturer	Yes	Yes	Yes	Yes
2	Technical & Engineering Capability	Yes	Yes	Yes	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes	Yes	Yes	Yes
4	Vendor Past Performance of Quoted Equipment	Yes	Yes	Yes	Yes
5	Availability of relevant tools and testing / calibration equipment	Yes	Yes	Yes	Yes
6	Compliance of Warranty as per tender	Yes	Yes	Yes	Yes
Remarks					
Eligible					

PART-III

Knock Down Criteria (Product Evaluation)

Specification Compliance / Evaluation Parameters					
	Excell 350 MCDSe	Surtron 400HP	hARTT 400	Maximum Smart C	
1	Model	ALSA	LED SPA	Kentamed	CLS Martin
2	Make	Italy	Italy	Bulgaria	Germany
3	Country of Origin	CE	CE	TUV NORD	CE
4	Certification				
5	Technical Specification	Yes	Yes	Yes	Yes
6	Microprocessor based electrosurgical unit for normal and under water cutting usages.	Yes	Yes	Yes	Yes
7	Automatic self-test function.	Yes	Yes	Yes	Yes

[Handwritten Signature]

[Handwritten Signature]

[Handwritten Signature]

8	Operation in radio frequency range.	22	Cutting, Coagulation, Spray and Blend	357 KHz	350-500KHz
9	Controls for cutting, coagulation, spray and blends.	22	Cutting, Coagulation, Spray and Blend	357 KHz	350-500KHz
10	Monopolar cutting power of 350 to 400 watts.	350 Watt	400 Watt	400 Watt	400 Watt
11	Bipolar cutting power of 140 watts.	140 Watt	400 Watt	300 Watt	120 Watt
12	Mono polar contact coagulation power of 220 to 250 Watts.	250 Watt	120 Watt	150	200 Watt
13	Bipolar coagulation power of 120 Watts.	120 Watt	120 Watt	150	250 Watt
14	Spray coagulation mode.	Yes	Yes	Yes	Yes
15	Different gradations of blending of cutting and coagulation power.	Yes	Yes	Yes	Yes
16	Digital display of all controls and set values of cutting and coagulation power.	Yes	Yes	Yes	Yes
17	Audio and visual alarms.	Yes	Yes	Yes	Yes
18	220V, 50 Hz.	Yes	Yes	Yes	Yes
19	ACCESSORIES				
20	Manipulator handle with cord	Offered	Offered	Offered	Offered
21	Bipolar forceps with cord.	Offered	Offered	Offered	Offered
22	Kafo electrode	Offered	Offered	Offered	Offered
23	Wire loop electrode	Offered	Offered	Offered	Offered
24	Surgical electrode, ball-shaped.	Offered	Offered	Offered	Offered
25	Needle electrode.	Offered	Offered	Offered	Offered
26	Attachment for monopolar coagulation.	Offered	Offered	Offered	Offered
27	Ball electrode	Offered	Offered	Offered	Offered
28	Bipolar coagulation forceps.	Offered	Offered	Offered	Offered
29	Reusable silicon patient plate.	Offered	Offered	Offered	Offered
30	Double paddle foot switch, explosion proof	Offered	Offered	Offered	Offered
31	Trolley with lockable anti-static castors	Local	Local	Offered	Local
32	COUNTRY OF ORIGIN: USA, EEC OR JAPAN	EEC	EEC	EEC	EEC
SPECIFICATION COMPLIANCE		Yes	No	No	No
TECHNICAL ELIGIBILITY OF THE PRODUCT		Yes	No	No	No
TECHNICAL ELIGIBILITY OF THE FIRM		Yes	Yes	Yes	Yes
Bid Status		Responsive	No-Responsive	Non-Responsive	Non-Responsive

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THQ HOSPITAL BUREWALA DISTRICT VEHARI
TECHNICAL EVALUATION FOR THE PURCHASE OF LARYNGOSCOPE
QTY-05, ESTIMATED COST Rs. 5000/-**

S# 7B- LARYNGOSCOPE

PART-I

Knock Down Criteria (Bid Evaluation)

Sr#	Evaluation Parameters	M/s Orient Medical
1	Complete Package / tender	Yes
2	Original Receipt of Tender	Yes
3	Affidavit from Bidder	Yes
4	Bid Security	Yes
5	Bid Validity	Yes
6	Delivery Period	Eligible
Remarks		

PART-II

Knock Down Criteria (Vendor Evaluation)

1	Valid Authorization Certificate from Manufacturer	Yes
2	Technical & Engineering Capability	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes
4	Vendor Past Performance of Quoted Equipment	Yes
5	Availability of relevant tools and testing / calibration equipment	Yes
6	Compliance of Warranty as per tender	Eligible
Remarks		

PART III

Knock Down Criteria (Product Evaluation)

Specification Compliance / Evaluation Parameters		
1	Model	Oreint
2	Make	Pak
3	Country of Origin	Pakistan







4	Certification		ISO
5	Technical Specification		
6	Machinosh type.		Yes
7	Blade set of 4 or 5.		Yes
8	Blade Sizes 1, 2, 3, 4 /5.		Yes
9	SS/ corrosion free Blades.		Yes
10	Dry Battery handle.		Yes
11	Blades of stainless steel with Integral light carrier.		Yes
12	Fiber optic light carrier.		Yes
13	Xenon illumination of light source.		Yes
14	Complete with batteries and carrying case.		Yes
15	Local Made Subject to Approval of Sample		Yes
		SPECIFICATION COMPLIANCE	
		TECHNICAL ELIGIBILITY OF THE PRODUCT	Yes
		TECHNICAL ELIGIBILITY OF THE FIRM	Yes
		Bid Status	Yes
			Responsive

Handwritten signature

Handwritten signature

Handwritten signature

Handwritten signature

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THQ HOSPITAL BUREWALA DISTRICT VEHARI**

**TECHNICAL EVALUATION FOR THE PURCHASE OF PULSE OXIMETER FOR PAEDS
QTY-01, ESTIMATED COST Rs. 150,000/-**

S# 8B - PULSE OXIMETER FOR PAEDS

PART-I

Knock Down Criteria (Bid Evaluation)

Sr#	Evaluation Parameters	Remarks
1	Complete Package / tender	No Firm Participated
2	Original Receipt of Tender	
3	Affidavit from Bidder	
4	Bid Security	
5	Bid Validity	
6	Delivery Period	

PART-II

Knock Down Criteria (Vendor Evaluation)

1	Valid Authorization Certificate from Manufacturer	
2	Technical & Engineering Capability	
3	Certificate from Manufacturer about after sale Service through agent or itself	
4	Vendor Past Performance of Quoted Equipment	
5	Availability of relevant tools and testing / calibration equipment	
6	Compliance of Warranty as per tender	

PART III

Knock Down Criteria (Product Evaluation)

Specification Compliance / Evaluation Parameters






**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THQ HOSPITAL BUREWALA DISTRICT VEHARI**

TECHNICAL EVALUATION FOR THE PURCHASE OF BABY INCUBATOR-INTENSIVE CARE INCUBATOR

QTY-02, ESTIMATED COST Rs. 15,00,000/-

Sr#9B - BABY INCUBATOR-INTENSIVE CARE INCUBATOR

PART-I

Knock Down Criteria (Bid Evaluation)

Sr#	Evaluation Parameters	M/s Sigma International	M/s Radiant Medical
1	Complete Package / tender	Yes	Yes
2	Original Receipt of Tender	Yes	Yes
3	Affidavit from Bidder	Yes	Yes
4	Bid Security	Yes	Yes
5	Bid Validity	Yes	Yes
6	Delivery Period	Yes	Yes
		Remarks	Eligible

PART-II

Knock Down Criteria (Vendor Evaluation)

1	Valid Authorization Certificate from Manufacturer	Yes	Yes
2	Technical & Engineering Capability	Yes	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes	Yes
4	Vendor Past Performance of Quoted Equipment	Yes	Yes
5	Availability of relevant tools and testing / calibration equipment	Yes	Yes
6	Compliance of Warranty as per tender	Yes	Yes
		Remarks	Eligible

PART III

Knock Down Criteria (Product Evaluation)

[Handwritten Signature]
Chaman = [Signature]

[Handwritten Signature]

Specification Compliance / Evaluation Parameters			
1	Model	Shelly	Arnelie
2	Make	TSE Medical	Ardo Medical
3	Country of Origin	CZECH REPUBLIC	Switzerland
4	Certification		CE
5	Technical Specification		
6	Primary Care Incubator Complexes	Yes	Yes
7	Microprocessor controlled unit with 5-inch LCD display of set Measure temperature & Air temperature control.	Yes	LCD 5"??
8	Temperature regulation range 28 C - 38 C +/- 1 C with uniform heat distribution over the bed surface.	20°C-37°C 34-37.9 Skin Uniform Heat	20°C-37°C 34-37 Skin Uniform heat
9	Replaceable bacteria filter for filtration of fresh air and oxygen.	Disinfectant	Disinfectant Yes
10	Bed surface 700 mm x 460 mm (approx.) with mattress	660mm x 4000 mm	1060mm x 680mm
11	Bed surface can be Trendelenburg and anti-Trendelenburg	Yes	Yes
12	Highly thermo stable acrylic glass or similar material double / Triple wall with four iris / elbow (steve) window and one side port		Yes
13	Humidification of the incubator air regulate able between 40-60 %	40-95%	0-90%
14	Oxygen flowmeter 0-15 liter/min	?	??
15	Noise Level: Equal or less than 45 dBA	Yes	45dBA
16	Alarm signals safety	Yes	Yes
17	The visual and audible battery powered alarm for power failure	?	Yes
18	Visual & Audible alarm for temperature excess (39 C)	?	Yes
19	A visual and audible alarm warns in case of fan failure	?	Yes
20	Test buttons to check the individual alarms or automatic test system	?	??
21	Accessories:		
22	Complete with temperature sensor	Offered	Offered
23	IV pole	Offered	Offered
24	Hand held pulse Oximeter built in / separate	Offered	8500 Nonin, USA

[Handwritten signature]

Chagan ~
Arjuna

[Handwritten signature]

25	Guard rail	Offered	Offered
26	Restraint straps	Offered	Offered
27	Country of Manufacturer: USA, EU AND JAPAN		
		SPECIFICATION COMPLIANCE	
		TECHNICAL ELIGIBILITY OF THE PRODUCT	No
		TECHNICAL ELIGIBILITY OF THE FIRM	No
		Bid Status	Yes
			non-responsive
			non-responsive

H. H. H.

Chiranjit

L. V. S.

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THQ HOSPITAL BUREWALA DISTRICT VEHARI**

**TECHNICAL EVALUATION FOR THE PURCHASE OF PHOTOTHERAPY UNIT
QTY-02, ESTIMATED COST Rs. 300,000/-**

S# 10B - PHOTOTHERAPY UNIT

PART-I

Knock Down Criteria (Bid Evaluation)

S#	Evaluation Parameters	Remarks
1	Complete Package / tender	No Firm Participated
2	Original Receipt of Tender	
3	Affidavit from Bidder	
4	Bid Security	
5	Bid Validity	
6	Delivery Period	

PART-II

Knock Down Criteria (Vendor Evaluation)

1	Valid Authorization Certificate from Manufacturer	
2	Technical & Engineering Capability	
3	Certificate from Manufacturer about after sale Service through agent or itself	
4	Vendor Past Performance of Quoted Equipment	
5	Availability of relevent tools and testing / calibration equipment	
6	Compliance of Warranty as per tender	

PART III

Knock Down Criteria (Product Evaluation)

Specification Compliance / Evaluation Parameters







**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THQ HOSPITAL BUREWALA DISTRICT VEHARI**

**TECHNICAL EVALUATION FOR THE PURCHASE OF FESS Instrument
(Qty-01) Estimated Cost. Rs. 100,000/- Each**

S# 118 FESS Instrument

PART-I

Knock Down Criteria (Bid Evaluation)

Sr#	Evaluation Parameters	M/s Noorani Surgical (Pvt) Ltd
1	Complete Package / tender	Yes
2	Original Receipt of Tender	Yes
3	Affidavit from Bidder	Yes
4	Bid Security	Yes
5	Bid Validity	Yes
6	Delivery Period	Yes
Remarks		Eligible

PART-II

Knock Down Criteria (Vendor Evaluation)

1	Valid Authorization Certificate from Manufacturer	Yes
2	Technical & Engineering Capability	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes
4	Vendor Past Performance of Quoted Equipment	Yes
5	Availability of relevant tools and testing / calibration equipment	Yes
6	Compliance of Warranty as per tender	Yes
Remarks		Eligible

PART III

Knock Down Criteria (Product Evaluation)







**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR UPGRADATION OF
THQ HOSPITAL BUREWALA DISTRICT VEHARI**

**TECHNICAL EVALUATION FOR THE PURCHASE OF XENON LIGHT SOURCE 300 WATT
QTY-1, ESTIMATED COST Rs. 100,000/-**

S# 12B - XENON LIGHT SOURCE 300 WATT

Knock Down Criteria (Bid Evaluation)

S#	Evaluation Parameters	Remarks
1	Complete Package / tender	No Firm Participated
2	Original Receipt of Tender	
3	Affidavit from Bidder	
4	Bid Security	
5	Bid Validity	
6	Delivery Period	

PART-II

Knock Down Criteria (Vendor Evaluation)

1	Valid Authorization Certificate from Manufacturer	
2	Technical & Engineering Capability	
3	Certificate from Manufacturer about after sale Service through agent or itself	
4	Vendor Past Performance of Quoted Equipment	
5	Availability of relevant tools and testing / calibration equipment	
6	Compliance of Warranty as per tender	

PART III

Knock Down Criteria (Product Evaluation)

Specification Compliance / Evaluation Parameters

--	--	--








**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THQ HOSPITAL BUREWALA DISTRICT VEHARI**

**TECHNICAL EVALUATION FOR THE PURCHASE OF DIAGNOSTIC VIDEO BRONCHOSCOPE
QTY-01, ESTIMATED COST Rs. 10,00,000/-**

S# 13B - DIAGNOSTIC VIDEO BRONCHOSCOPE

PART-I

Knock Down Criteria (Bid Evaluation)






S#	Evaluation Parameters	M/s Mediland
1	Complete Package / tender	Yes
2	Original Receipt of Tender	Yes
3	Affidavit from Bidder	Yes
4	Bid Security	Yes
5	Bid Validity	Yes
6	Delivery Period	Eligible
PART-II		
Knock Down Criteria (Vendor Evaluation)		

1	Valid Authorization Certificate from Manufacturer	Yes
2	Technical & Engineering Capability	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes
4	Vendor Past Performance of Quoted Equipment	Yes
5	Availability of relevant tools and testing / calibration equipment	Yes
6	Compliance of Warranty as per tender	Eligible

PART III

Knock Down Criteria (Product Evaluation)

Specification Compliance / Evaluation Parameters

1	Model	EB15-J10
2	Make	Pentax Medical
3	Country of Origin	Japan
4	Certification	JIS
5	Technical Specification	
6	Brilliant, high-definition HD+ images for a clear and more detailed visualization.	Yes
7	Greatest possible insertion tube flexibility and excellent maneuverability due to a very small distal tip bending radius.	Yes
8	Digital signal processing design to optimize Color and contrast.	Yes
9	Full screen crisp, clear image	Yes
10	IEE digital image enhanced technology I-Scan with surface, contrast and tone enhancement to improve vessel architecture evaluation.	Yes
11	Lightweight control body for superior efficiency, accuracy and control	Yes
12	Ergonomically positioned.	Yes
13	Suction valve and angulations levers for ease of operation.	Yes
14	Customizable buttons for individual usability and operation with one hand	Yes
15	operation with one hand	Yes
16	It Should be capable for FVT (Full Vision Technology)	Yes
17	180-degree rotatable Anti-loop connector for scope handling.	Yes
18	Distal end should be insulated for use with electrosurgical devices to provide increased patient safety.	Yes
19	2x magnification / Zoom.	Yes
20	Direction of View Forward	Forward
21	Insertion Tube Diameter 5.2 mm	5.2mm
22	Distal end diameter 5.5 mm	5.4mm
23	Channel Diameter 2.0 mm	2mm
24	Working Length 600 mm	600mm
25	Field of View 120°	120
26	Angulations (Up / Down 210 /130	210/130
27	Total Length 850 mm	Yes

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THQ HOSPITAL BUREWALA DISTRICT VEHARI
TECHNICAL EVALUATION FOR THE PURCHASE OF LARYNGOSCOPE FOR ENT
QTY-01, ESTIMATED COST Rs. 400,000/-**

S# 14B - LARYNGOSCOPE FOR ENT

PART-I

Knock Down Criteria (Bid Evaluation)

S#	Evaluation Parameters	M/s Sigma International	M/s Mediland
1	Complete Package / tender	Yes	Yes
2	Original Receipt of Tender	Yes	Yes
3	Affidavit from Bidder	Yes	Yes
4	Bid Security	Yes	Yes
5	Bid Validity	Yes	Yes
6	Delivery Period	Yes	Yes
		Remarks	Eligible

PART-II

Knock Down Criteria (Vendor Evaluation)

1	Valid Authorization Certificate from Manufacturer	Yes	Yes
2	Technical & Engineering Capability	Yes	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes	Yes
4	Vendor Past Performance of Quoted Equipment	Yes	Yes
5	Availability of relevant tools and testing / calibration equipment	Yes	Yes
6	Compliance of Warranty as per tender	Yes	Yes
		Remarks	Eligible

PART III

Knock Down Criteria (Product Evaluation)

Remarks	Eligible	Eligible
---------	----------	----------

[Handwritten signatures and initials]

[Signature]

[Signature]

[Signature]

Specification Compliance / Evaluation Parameters

Specification Compliance / Evaluation Parameters		Naso-pharyngo-laryngo Scope Entered	VNL9-CP Pentax Medical Japan JIC
1	Model	Naso-pharyngo-laryngo Scope Entered	VNL9-CP Pentax Medical Japan JIC
2	Make	Netherlands	Japan JIC
3	Country of Origin		
4	Certification		
5	Technical Specification		
6	CMOS / CCD Video-Rhino-Laryngoscope.		CCD
7	Direction of view: 0°		0
8	Angle of view: 75 to 85° or m		90
9	Deflection: up down: 120-140°		120/120
10	Working length: 30 cm		300mm
11	Oulter Diameter: 3-3.7 mm		3mm
12	Light Source:		100 Watt LED
13	LED / Xenon 100 watt or better Light Source		Integrated
14	Light Built in with Monitor or Separately both are acceptable		
15	Monitor:		21" Endovue LED HD
16	Model		NDS Surgical
17	Make		USA
18	Country of Origin		21"
19	Medical Graded Monitor 19" or better		
20	Video Processor:		CP-1000
21	Model		Pentax Medical
22	Make		Japan
23	Country of Origin		Integrated
24	Video Processor Resolution 720 x 1280 or better Built in or Separate		Yes
25	Output Ports: - DVI / RGB / HDMI		Yes
26	case for storing Video Fibero Laryngoscope		Yes

No Data Sheet Provided

Handwritten signature
Onigane / anigab

Handwritten signature

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THQ HOSPITAL BUREWALA DISTRICT VEHARI**

**TECHNICAL EVALUATION FOR THE PURCHASE OF Stretcher Trolley (High Low Type)
(Qty-04) Estimated Cost. Rs. 35,000/- Each**

Sr# 15B Stretcher Trolley (High Low Type)

PART-I

Knock Down Criteria (Bid Evaluation)

Sr#	Evaluation Parameters	M/s Thankful Agencies	M/s Orient Medical	M/s Noorani Surgical (Pvt) Ltd	M/s Pieces Engineering & GOS
1	Complete Package / tender	Yes	Yes	Yes	Yes
2	Original Receipt of Tender	Yes	Yes	Yes	Yes
3	Affidavit from Bidder	Yes	Yes	Yes	Yes
4	Bid Security	Yes	Yes	Yes	No
5	Bid Validity	Yes	Yes	Yes	Yes
6	Delivery Period	Yes	Yes	Yes	Yes
	Remarks	Eligible	Eligible	Eligible	Not-Eligible

PART-II

Knock Down Criteria (Vendor Evaluation)

1	Valid Authorization Certificate from Manufacturer	Yes	Yes	Yes	Yes
2	Technical & Engineering Capability	Yes	Yes	Yes	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes	Yes	Yes	Yes
4	Vendor Past Performance of Quoted Equipment	Yes	Yes	Yes	Yes
5	Availability of relevant tools and testing / calibration equipment	Yes	Yes	Yes	Yes
6	Compliance of Warranty as per tender	Yes	Yes	Yes	Yes
	Remarks	Eligible	Eligible	Eligible	Eligible

PART-III

Knock Down Criteria (Product Evaluation)








Specification Compliance / Evaluation Parameters

1	Model	Local	Local	Local	Local
2	Make	Thankful Agencies	Local	Local	Local
3	Country of Origin	Pakistan	Pakistan	Pakistan	Pakistan
4	Technical Specification				
5	Stretcher trolley (high low type)	Yes	Yes	Yes	Yes
6	Complete SS Structure	Yes	Yes	Yes	Yes
7	2 sections Patient Shifting Trolley.	Yes	Yes	Yes	Yes
8	Oxygen Bottle Holder Wire basket	Yes	Yes	Yes	Yes
9	Bumper at all corner to protect trolley	Yes	Yes	Yes	Yes
10	Patient Safe Working Load capacity min 150 kg or more.	Yes	Yes	Yes	Yes
11	Local Made Subject to approval of sample				
SPECIFICATION COMPLIANCE		No	No	No	No
TECHNICAL ELIGIBILITY OF THE PRODUCT		Yes	No	No	Yes
TECHNICAL ELIGIBILITY OF THE FIRM		Yes	Yes	Yes	Yes
Bid Status		Non-Responsive	Non-Responsive	Non-Responsive	Non-Responsive

Zameer-ajaz

Chahar

1-1-2024

Sample is not as per stated spec

Sample not provided

Sample is not provided

Sample is not as per required spec's

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THQ HOSPITAL BUREWALA DISTRICT VEHARI
TECHNICAL EVALUATION FOR THE PURCHASE OF Ward Screen with Curtain
(Qty-08) Estimated Cost. Rs. 4,000/- Each**

Sr# 168 Ward Screen with Curtain

PART-I

Knock Down Criteria (Bid Evaluation)

Sr#	Evaluation Parameters	M/s Pieces Engineering & GOS	M/s Orient Medical	M/s Noorani Surgical (Pvt) Ltd
1	Complete Package / tender	Yes	Yes	Yes
2	Original Receipt of Tender	Yes	Yes	Yes
3	Affidavit from Bidder	Yes	Yes	Yes
4	Bid Security	No	Yes	Yes
5	Bid Validity	Yes	Yes	Yes
6	Delivery Period	Yes	Yes	Yes
Remarks		Not-Eligible	Eligible	Eligible

PART-II

Knock Down Criteria (Vendor Evaluation)

1	Valid Authorization Certificate from Manufacturer	Yes	Yes	Yes
2	Technical & Engineering Capability	Yes	Yes	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes	Yes	Yes
4	Vendor Past Performance of Quoted Equipment	Yes	Yes	Yes
5	Availability of relevant tools and testing / calibration equipment	Yes	Yes	Yes
6	Compliance of Warranty as per tender	Yes	Yes	Yes
Remarks		Eligible	Eligible	Eligible

PART III

Knock Down Criteria (Product Evaluation)








Specification Compliance / Evaluation Parameters

1	Model	Local	Local	Local
2	Make	Local	Local	Local
3	Country of Origin	Pakistan	Pakistan	Pakistan
4	Technical Specification			
5	Frame in four (04) parts constructed from fully plastic coated, MS hollow tubular pipe 1" dia.	✓ Yes	Yes	Yes
6	All frames fixed together with ABS universal joint for easy movement of each part	✓ Yes	Yes	Yes
7	Curtain fitted with spring loaded rods mobile on six castors of 75mm dia	✓ Yes	Yes	Yes
8	Wearing washable polyester cloth in variety of color. <i>Rule cloth.</i>	✓ Yes	Yes	Yes
9	All metal parts to be chemically degreased / derusted through phosphate treatment OR sandblasting and covered with electrostatic polyester powder coating	✓ Yes	Yes	Yes
10	Local Made Subject to Approval of Sample			
		SPECIFICATION COMPLIANCE		
		TECHNICAL ELIGIBILITY OF THE PRODUCT	Yes	No
		TECHNICAL ELIGIBILITY OF THE FIRM	Yes	Yes
		Bid Status	Non-Responsive	Non-Responsive

Handwritten signature

Handwritten signature

Handwritten signature

Handwritten signature

Handwritten signature

Non-Responsive
Simple not provided

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THQ HOSPITAL BUREWALA DISTRICT VEHARI**

**TECHNICAL EVALUATION FOR THE PURCHASE OF SS Basin with SS Stand
(Qty-08) Estimated Cost. Rs. 6,000/- Each**

Sr# 17B - SS Basin with SS Stand

PART-I

Knock Down Criteria (Bid Evaluation)

Sr#	Evaluation Parameters	M/s Thankful Agencies	M/s Orient Medical	M/s Noorani Surgical (Pvt) Ltd
1	Complete Package / tender	Yes	Yes	Yes
2	Original Receipt of Tender	Yes	Yes	Yes
3	Affidavit from Bidder	Yes	Yes	Yes
4	Bid Security	Yes	Yes	Yes
5	Bid Validity	Yes	Yes	Yes
6	Delivery Period	Yes	Yes	Yes
	Remarks	Eligible	Eligible	Eligible

PART-II

Knock Down Criteria (Vendor Evaluation)

1	Valid Authorization Certificate from Manufacturer	Yes	Yes	Yes
2	Technical & Engineering Capability	Yes	Yes	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes	Yes	Yes
4	Vendor Past Performance of Quoted Equipment	Yes	Yes	Yes
5	Availability of relevant tools and testing / calibration equipment	Yes	Yes	Yes
6	Compliance of Warranty as per tender	Yes	Yes	Yes
	Remarks	Eligible	Eligible	Eligible

PART III

Knock Down Criteria (Product Evaluation)








Specification Compliance / Evaluation Parameters

1	Model	Local	Local	Local
2	Make	Thankful Agencies	Local	Local
3	Country of Origin	Pakistan	Pakistan	Pakistan
4	Technical Specification			
5	Complete Stainless Steel Structure	Yes	Yes	Yes
6	Local Made Subject to Approval of Sample			
SPECIFICATION COMPLIANCE		No	Yes	No
TECHNICAL ELIGIBILITY OF THE PRODUCT		No	Yes	No
TECHNICAL ELIGIBILITY OF THE FIRM		Yes	Yes	Yes
Bid Status		Non-Responsive	Responsive	Non-Responsive

Signature
Signature

Signature

Non-Responsive
Sample not provided

Signature

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THQ HOSPITAL BUREWALA DISTRICT VEHARI**

**TECHNICAL EVALUATION FOR THE PURCHASE OF Dressing Trolley
(Qty-04) Estimated Cost. Rs. 25,000/- Each**

Sr#18B - Dressing Trolley

PART-I

Knock Down Criteria (Bid Evaluation)

Sr#	Evaluation Parameters	M/s Pieces Engineering & GOS	M/s Orient Medical	M/s Noorani Surgical (Pvt) Ltd
1	Complete Package / tender	Yes	Yes	Yes
2	Original Receipt of Tender	Yes	Yes	Yes
3	Affidavit from Bidder	Yes	Yes	Yes
4	Bid Security	No	Yes	Yes
5	Bid Validity	Yes	Yes	Yes
6	Delivery Period	Yes	Yes	Yes
	Remarks	Not-Eligible	Eligible	Eligible

PART-II

Knock Down Criteria (Vendor Evaluation)

1	Valid Authorization Certificate from Manufacturer	Yes	Yes	Yes
2	Technical & Engineering Capability	Yes	Yes	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes	Yes	Yes
4	Vendor Past Performance of Quoted Equipment	Yes	Yes	Yes
5	Availability of relevant tools and testing / calibration equipment	Yes	Yes	Yes
6	Compliance of Warranty as per tender	Yes	Yes	Yes
	Remarks	Eligible	Eligible	Eligible

PART III

Knock Down Criteria (Product Evaluation)

Specification Compliance / Evaluation Parameters

1	Model	Local	Local	Local
2	Make	Local	Local	Local
3	Country of Origin	Pakistan	Pakistan	Pakistan
4	Technical Specification			
5	Mobile type of Complete stainless-steel rust free	Yes	✓ Yes	Yes
6	4 leg frames made of Stainless-steel pipe 16 SWG 1"x1" all made of stainless steel	Yes	✓ Yes	Yes
7	Tubular frame forms a safety rail, no sharp edges.	Yes	✓ Yes	Yes
8	Two shelves surface ground to reduce glare and resistant to disinfectants.	Yes	✓ Yes	Yes
9	Trolley, 800mm high, fitted above and below with frames and shelf 20 SWG made, size 24" x 18", to accept two sterilization trays.	Yes	✓ Yes	Yes
10	1xSS ring for bowl 10" with SS bowl and 1xSS ring for bucket with SS bucket.	Yes	✓ Yes	Yes
11	Bottle rack at top shelf.	Yes	✓ Yes	Yes
12	Four 125mm swivel castors, electrically conductive. Dimensions are approximate.	Yes	✓ Yes	Yes
13	Local Made Subject to Approval of Sample			
SPECIFICATION COMPLIANCE				
TECHNICAL ELIGIBILITY OF THE PRODUCT		Yes	Yes	No
TECHNICAL ELIGIBILITY OF THE FIRM		Yes	Yes	No
Bid Status		Non-Responsive	Responsive	Non-Responsive

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

Non-Responsive
Sample not provided

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THQ HOSPITAL BUREWALA DISTRICT VEHARI
TECHNICAL EVALUATION FOR THE PURCHASE OF Medicine Trolley
(Qty-08) Estimated Cost. Rs. 15,000/- Each**

Sr#19B - Medicine Trolley

PART-I

Knock Down Criteria (Bid Evaluation)

Sr#	Evaluation Parameters	M/s Thankful Agencies	M/s Noorani Surgical (Pvt) Ltd	M/s Orient Medical	M/s Pieces Engineering & GOS
1	Complete Package / tender	Yes	Yes	Yes	Yes
2	Original Receipt of Tender	Yes	Yes	Yes	Yes
3	Affidavit from Bidder	Yes	Yes	Yes	Yes
4	Bid Security	Yes	Yes	Yes	No
5	Bid Validity	Yes	Yes	Yes	Yes
6	Delivery Period	Yes	Yes	Yes	Yes
	Remarks	Eligible	Eligible	Eligible	Not-Eligible

PART-II

Knock Down Criteria (Vendor Evaluation)

1	Valid Authorization Certificate from Manufacturer	Yes	Yes	Yes	Yes
2	Technical & Engineering Capability	Yes	Yes	Yes	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes	Yes	Yes	Yes
4	Vendor Past Performance of Quoted Equipment	Yes	Yes	Yes	Yes
5	Availability of relevant tools and testing / calibration equipment	Yes	Yes	Yes	Yes
6	Compliance of Warranty as per tender	Yes	Yes	Yes	Yes
	Remarks	Eligible	Eligible	Eligible	Eligible

PART III

Knock Down Criteria (Product Evaluation)

Specification Compliance / Evaluation Parameters

Chaudhary Anwar
Anwar

1	Model	Local	Local	Local	Local
2	Make	Thankful Agencies	Local	Local	Local
3	Country of Origin	Pakistan	Pakistan	Pakistan	Pakistan
4	Technical Specification				
5	Fiber Structure for the storage of medication by ward nursing staff	Yes	Yes	Yes	Yes
6	Mounted on approx. 125mm castors.	Yes	Yes	Yes	Yes
7	The approximate internal dimensions of the medicine container shall be 80 w x50 d x30 h cm. approx.	Yes	Yes	68x57x32	Yes
8	Lockable with heavy-duty dead lock	Yes	Yes	Yes	Yes
9	Shall be fitted with the following accessories:				
10	4 tier rack for holding medications	Yes	Yes	Yes	Yes
11	Ampoule holding rack.	Yes	Yes	Yes	Yes
12	Side mounted fold-down writing flap.	Yes	Yes	Yes	Yes
13	Different compartments for medicines with labeling.	Yes	Yes	Yes	Yes
14	Lower shelf mounted under medicine container.	Yes	Yes	Yes	Yes
15	Unit shall be supplied with a security clamp for wall fixing at the appropriate location, thus ensuring that the trolley is secure when not in use	Yes	Yes	Yes	Yes
16	Local Made Subject to Approval of Sample				
SPECIFICATION COMPLIANCE		N/A	N/A	N/A	N/A
TECHNICAL ELIGIBILITY OF THE PRODUCT		N/A	N/A	N/A	N/A
TECHNICAL ELIGIBILITY OF THE FIRM		Yes	Yes	Yes	Yes
Bid Status		Non-Responsive	Non-Responsive	Non-Responsive	Non-Responsive

[Handwritten signatures]

Sample is not as per spec 2

Sample not provided

Sample is not as per required specs

Sample not provided

[Handwritten mark]

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THQ HOSPITAL BUREWALA DISTRICT VEHARI
TECHNICAL EVALUATION FOR THE PURCHASE OF Instrument Cabinet
(Qty-20) Estimated Cost. Rs. 20,000/- Each**

Sr# 208 - Instrument Cabinet

PART-I

Knock Down Criteria (Bid Evaluation)

Sr#	Evaluation Parameters	M/s Pieces Engineering & GOS	M/s Orient Medical	M/s Noorani Surgical (Pvt) Ltd
1	Complete Package / tender	Yes	Yes	Yes
2	Original Receipt of Tender	Yes	Yes	Yes
3	Affidavit from Bidder	Yes	Yes	Yes
4	Bid Security	No	Yes	Yes
5	Bid Validity	Yes	Yes	Yes
6	Delivery Period	Yes	Yes	Yes
Remarks		Not-Eligible	Eligible	Eligible

PART-II

Knock Down Criteria (Vendor Evaluation)

1	Valid Authorization Certificate from Manufacturer	Yes	Yes	Yes
2	Technical & Engineering Capability	Yes	Yes	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes	Yes	Yes
4	Vendor Past Performance of Quoted Equipment	Yes	Yes	Yes
5	Availability of relevant tools and testing / calibration equipment	Yes	Yes	Yes
6	Compliance of Warranty as per tender	Yes	Yes	Yes
Remarks		Eligible	Eligible	Eligible

PART III

Knock Down Criteria (Product Evaluation)

Specification Compliance / Evaluation Parameters

Chaudhary Anwar

Arif

L-T Team

Arif

1	Model		Local	Local	Local
2	Make		Local	Local	Local
3	Country of Origin		Pakistan	Pakistan	Pakistan
4	Technical Specification				
5	Instrument and medicine cabinet of mild steel with EPC finish.		Yes	Yes	Yes
6	Floor standing with double locking.		Yes	Yes	Yes
7	Made of 20G steel		Yes	Yes	Yes
8	Glazed doors and 6 glass shelves.		Yes	Yes	Yes
9	Size of 150 X 90 X 40cm.		Yes	Yes	Yes
10	OR		Yes	Yes	Yes
11	Size: 6 ft x 3 ft x 1 ft 6 inch (H x D x W)		Yes	Yes	Yes
12	All made of mild steel 22 gauge with 4 shelves with lockable doors including glass finished, silver hammer colour		Yes	Yes	Yes
13	Local Made Subject to Approval of Sample				
		SPECIFICATION COMPLIANCE			
		TECHNICAL ELIGIBILITY OF THE PRODUCT	No	Yes	No
		TECHNICAL ELIGIBILITY OF THE FIRM	No	Yes	No
		Bid Status	Non-Responsive	Responsive	Non-Responsive

Handwritten signature

Handwritten signature

Handwritten signature

Handwritten signature

Handwritten signature

Non-Responsive

Non-Responsive

Sample not provided


**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THQ HOSPITAL BUREWALA DISTRICT VEHARI
TECHNICAL EVALUATION FOR THE PURCHASE OF BP APPARATUS WITH STAND MODEL
QTY-20, ESTIMATED COST Rs. 12000/- EACH**

Sr# 21 - BP APPARATUS WITH STAND MODEL


Knock Down Criteria (Bid Evaluation)				
Sr#	Evaluation Parameters	M/s Orient Medical	M/s Med Impex	M/s Noorani Surgical
1	Complete Package / tender	Yes	Yes	Yes
2	Original Receipt of Tender	Yes	Yes	Yes
3	Affidavit from Bidder	Yes	Yes	Yes
4	Bid Security	Yes	Yes	Yes
5		Yes	Yes	Yes
6	Delivery Period	Yes	Yes	Yes
	Remarks	Eligible	Eligible	Eligible

PART-II Knock Down Criteria (Vendor Evaluation)				
1	Valid Authorization Certificate from Manufacturer	Yes	Yes	No
2	Technical & Engineering Capability	Yes	Yes	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes	Yes	Yes
4	Vendor Past Performance of Quoted Equipment	Yes	Yes	Yes
5	Availability of relevent tools and testing / calibration equipment	Yes	Yes	Yes
6	Compliance of Warranty as per tender	Yes	Yes	Yes
	Remarks	Eligible	Eligible	Not-Eligible

PART III Knock Down Criteria (Product Evaluation)				
Specification Compliance / Evaluation Parameters				
1	Model	Orient	M-120	??





**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THQ HOSPITAL BUREWALA DISTRICT VEHARI
TECHNICAL EVALUATION FOR THE PURCHASE OF Instrument Tray
(Qty-16) Estimated Cost. Rs. 1,000/- Each**

S# 22B - Instrument Tray

**PART-I
Knock Down Criteria (Bid Evaluation)**

Sr#	Evaluation Parameters	M/s Noorani Surgical (Pvt) Ltd	M/s Orient Medical
1	Complete Package / tender	Yes	Yes
2	Original Receipt of Tender	Yes	Yes
3	Affidavit from Bidder	Yes	Yes
4	Bid Security	Yes	Yes
5	Bid Validity	Yes	Yes
6	Delivery Period	Eligible	Eligible

PART-II

Knock Down Criteria (Vendor Evaluation)

1	Valid Authorization Certificate from Manufacturer	Yes	Yes
2	Technical & Engineering Capability	Yes	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes	Yes
4	Vendor Past Performance of Quoted Equipment	Yes	Yes
5	Availability of relevant tools and testing / calibration equipment	Yes	Yes
6	Compliance of Warranty as per tender	Eligible	Eligible

PART-III

Knock Down Criteria (Product Evaluation)

Specification Compliance / Evaluation Parameters	Local	Local
1 Model	Yes	Yes

Chaudhary

2	Make	Local	Local
3	Country of Origin	Pakistan	Pakistan
4	Technical Specification		
5	Tray Instrument 300x175x40mm with cover	Yes	Yes
6	Local Made Subject to Approval of Sample		
SPECIFICATION COMPLIANCE			
TECHNICAL ELIGIBILITY OF THE PRODUCT		No	Yes
TECHNICAL ELIGIBILITY OF THE FIRM		No	Yes
Bid Status		Not Responsive to sample not provided	Responsive








**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR UPGRADATION OF
THQ HOSPITAL BUREWALA DISTRICT VEHARI**

**TECHNICAL EVALUATION FOR THE PURCHASE OF OXYGEN CYLINDER LARGE SIZE
QTY-10, ESTIMATED COST Rs. 20,000/- EACH**

S# 238 - OXYGEN CYLINDER LARGE SIZE

Knock Down Criteria (Bid Evaluation)

S#	Evaluation Parameters	M/s Noorani Surgical
1	Complete Package / tender	Yes
2	Original Receipt of Tender	Yes
3	Affidavit from Bidder	Yes
4	Bid Security	Yes
5	Bid Validity	Yes
6	Delivery Period	Yes
Remarks		Eligible

PART-II



Knock Down Criteria (Vendor Evaluation)


1	Valid Authorization Certificate from Manufacturer	Yes
2	Technical & Engineering Capability	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes
4	Vendor Past Performance of Quoted Equipment	Yes
5	Availability of relevant tools and testing / calibration equipment	Yes
6	Compliance of Warranty as per tender	Yes
Remarks		Eligible

PART III

Knock Down Criteria (Product Evaluation)

Specification Compliance / Evaluation Parameters		Euro
1	Model	

2	Make		Germany
3	Country of Origin		??
4	Technical Specification		
5	240 Ctt Medical Cylinder		Yes
6	Nominal Capacity: 6800 ltrs/ 240cft		Yes
7	Nominal Pressure: 137 Bar		Yes
8	Water Capacity: 47 Ltr approx.		Yes
9	Bull nose valve		Yes
10	A handle wheel with transport safety cap shall be supplied with the cylinder.		Yes
11	Duly tested by renowned/ authorized company.		Yes
12	BS specifications and Color code standard.		Yes
13	Country of Origin: USA, EEC, JAPAN		EEC
		SPECIFICATION COMPLIANCE	Yes
		TECHNICAL ELIGIBILITY OF THE PRODUCT	Yes
		TECHNICAL ELIGIBILITY OF THE FIRM	Yes
		Bid Status	Responsive

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR UPGRADATION OF
THQ HOSPITAL BUREWALA DISTRICT VEHARI**

**TECHNICAL EVALUATION FOR THE PURCHASE OF OXYGEN CYLINDER MEDIUM SIZE
QTY-20, ESTIMATED COST Rs. 15000/- EACH**

Sr# 24B - OXYGEN CYLINDER MEDIUM SIZE

Knock Down Criteria (Bid Evaluation)

Sr#	Evaluation Parameters	M/s Noorani Surgical
1	Complete Package / tender	Yes
2	Original Receipt of Tender	Yes
3	Affidavit from Bidder	Yes
4	Bid Security	Yes
5	Bid Validity	Yes
6	Delivery Period	Yes
Remarks		Eligible

PART-II

Knock Down Criteria (Vendor Evaluation)

1	Valid Authorization Certificate from Manufacturer	Yes
2	Technical & Engineering Capability	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes
4	Vendor Past Performance of Quoted Equipment	Yes
5	Availability of relevant tools and testing / calibration equipment	Yes
6	Compliance of Warranty as per tender	Yes
Remarks		Eligible

PART III

Knock Down Criteria (Product Evaluation)

Specification Compliance / Evaluation Parameters

1	Model	Euro
---	-------	------

2	Make	Germany
3	Country of Origin	??
4	Technical Specification	
5	48 Ct Medical Cylinder	Yes
6	Bull nose valve	Yes
7	A handle wheel with transport safety cap shall be supplied with the cylinder.	Yes
8	Duly tested by renowned/ authorized company.	Yes
9	BS specifications and Color code standard.	Yes
10	Country of Origin: USA, EEC, JAPAN	Germany
	TECHNICAL ELIGIBILITY OF THE PRODUCT	Yes
	TECHNICAL ELIGIBILITY OF THE FIRM	Yes
	Bid Status	Responsive

Expert
anjib

on 12/11/2

[Signature]

[Signature]

TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR UPGRADATION OF

THQ HOSPITAL BUREWALA DISTRICT VEHARI

TECHNICAL EVALUATION FOR THE PURCHASE OF OXYGEN REGULATOR

QTY-50, ESTIMATED COST Rs. 1500/- EACH

S# 25B - OXYGEN REGULATOR

Knock Down Criteria (Bid Evaluation)

S#	Evaluation Parameters	Remarks
1	Complete Package / tender	No Firm Participated
2	Original Receipt of Tender	
3	Affidavit from Bidder	
4	Bid Security	
5	Bid Validity	
6	Delivery Period	

PART-II

Knock Down Criteria (Vendor Evaluation)

1	Valid Authorization Certificate from Manufacturer
2	Technical & Engineering Capability
3	Certificate from Manufacturer about after sale Service through agent or itself
4	Vendor Past Performance of Quoted Equipment
5	Availability of relevant tools and testing / calibration equipment
6	Compliance of Warranty as per tender

Remarks

PART III

Knock Down Criteria (Product Evaluation)

Specification Compliance / Evaluation Parameters

1	Model
---	-------

Handwritten signature

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THQ HOSPITAL BUREWALA DISTRICT VEHARI
TECHNICAL EVALUATION FOR THE PURCHASE OF Linen Trunk
(Qty-04) Estimated Cost. Rs. 8,000/- Each**

Sr# 26B - Linen Trunk

PART-I

Knock Down Criteria (Bid Evaluation)

Sr#	Evaluation Parameters	M/s Noorani Surgical (Pvt) Ltd
1	Complete Package / tender	Yes
2	Original Receipt of Tender	Yes
3	Affidavit from Bidder	Yes
4	Bid Security	Yes
5	Bid Validity	Yes
6	Delivery Period	Yes
Remarks		Eligible

PART-II

Knock Down Criteria (Vendor Evaluation)

1	Valid Authorization Certificate from Manufacturer	Yes
2	Technical & Engineering Capability	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes
4	Vendor Past Performance of Quoted Equipment	Yes
5	Availability of relevant tools and testing / calibration equipment	Yes
6	Compliance of Warranty as per tender	Yes
Remarks		Eligible

PART III

Knock Down Criteria (Product Evaluation)

Specification Compliance / Evaluation Parameters	Model	Local
1		

Shamir /

2	Make		Local
3	Country of Origin		Pakistan
4	Technical Specification		
5	Complete SS Structure		
6	Set of two linen receptacles with one spare canvas bag		Yes
7	Mounted on 4 rubber wheels dia 200 mm (2 fixed and 2 swivels) arranged at the corners.		Yes
8	Dimensions: L: 1030 mm, D:630 mm, H:710 mm approx		Yes
9	Local Made Subject to Approval of Sample		
		SPECIFICATION COMPLIANCE	No
		TECHNICAL ELIGIBILITY OF THE PRODUCT	No
		TECHNICAL ELIGIBILITY OF THE FIRM	Yes
		Bid Status	Non-Responsive

Signature
Signature

Signature

Signature

Signature

Sample not provided

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR UPGRADATION OF
THQ HOSPITAL BUREWALA DISTRICT VEHARI**

**TECHNICAL EVALUATION FOR THE PURCHASE OF WEIGHT SCALE
QTY-08, ESTIMATED COST Rs. 2000/- EACH**

S# 27B - WEIGHT SCALE

Knock Down Criteria (Bid Evaluation)

Sr#	Evaluation Parameters	Remarks
1	Complete Package / tender	No Firm Participated
2	Original Receipt of Tender	
3	Affidavit from Bidder	
4	Bid Security	
5	Bid Validity	
6	Delivery Period	

PART-II

Knock Down Criteria (Vendor Evaluation)

1	Valid Authorization Certificate from Manufacturer	
2	Technical & Engineering Capability	
3	Certificate from Manufacturer about after sale Service through agent or itself	
4	Vendor Past Performance of Quoted Equipment	
5	Availability of relevant tools and testing / calibration equipment	
6	Compliance of Warranty as per tender	Remarks

PART III

Knock Down Criteria (Product Evaluation)

Specification Compliance / Evaluation Parameters

1	Model	
---	-------	--







**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THQ HOSPITAL BUREWALA DISTRICT VEHARI**

**TECHNICAL EVALUATION FOR THE PURCHASE OF BABY WARMER WITH INTEGRATED BUILT-IN WEIGHT SCALE
QTY-01, ESTIMATED COST Rs.10,00,000/-**

Sr# 29B - BABY WARMER WITH INTEGRATED BUILT-IN WEIGHT SCALE

**PART-I
Knock Down Criteria (Bid Evaluation)**

Sr#	Evaluation Parameters	M/s Radiant Medical
1	Complete Package / tender	Yes
2	Original Receipt of Tender	Yes
3	Affidavit from Bidder	Yes
4	Bid Security	Yes
5	Bid Validity	Yes
6	Delivery Period	Eligible
PART-II		
Knock Down Criteria (Vendor Evaluation)		

1	Valid Authorization Certificate from Manufacturer	Yes
2	Technical & Engineering Capability	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes
4	Vendor Past Performance of Quoted Equipment	Yes
5	Availability of relevant tools and testing / calibration equipment	Yes
6	Compliance of Warranty as per tender	Eligible

**PART III
Knock Down Criteria (Product Evaluation)
Specification Compliance / Evaluation Parameters**








1	Model	Americ e-Servo
2	Make	Ardo Medical
3	Country of Origin	Switzerland
4	Certification	CE
5	Technical Specification	
6	Baby resuscitation Trolley with warming system,	Integrated warming
7	Microprocessor controlled heating system,	Microprocessor
8	Open intensive care system for pre-mature and newborn	Yes
9	Mobile with antistatic castor Lockable	Mobile with antistatic castors
10	Bumper guard	
11	Manual heat output control: 0% to 100%	0-100%
12	Skin and Manual temperature control settings	Yes
13	Heating power / source : 500 W Quartz/Ceramic	500 Watt Quartz
14	Pivot arm technology for heating,	Pivot arm
15	Display range of temperature: LED / LCD	LED
16	Selection for operating modes: Skin or Manual	Both
17	Head can be moved in both directions allowing X-Ray procedure without moving the baby.	Yes
18	Integrated baby bed 700 x 450 mm approx. with secured Plexiglas side panels, foldable down, with grid for X-ray.	1200mm x 75000
19	Integrated observation lamp	?? Yes
20	Manual bed inclination	10 Degree Both
21	ALARM	
22	Audio and visual alarms for Power failure,	Yes
23	Skin Temperature deviations,	Yes
24	High Temperature,	Yes
25	Skin probe defective / unplugged,	Yes
26	Accessories :	?? Yes
27	Combined O2 humidifier with venturi suction, complete including flow meter and suction bottle dedicated to neonates.	?? Yes

[Handwritten signatures and initials in blue ink, including 'Claus', 'D. W. ...', and 'L. H. ...']

28	Corrugated tube for O2 humidifier and O2 connection hose.	Offered	Yes
29	1 x O2 Cylinder with fittings.	Offered	
30	Skin probe (reusable)	Offered	
31	IV pole	Offered	
32	Complete with O2 hood		
33	OPTIONAL:	Offered	
34	Resuscitation device for babies	Offered	
35	Breathe delivery system to deliver CPAP and PEEP	Offered	
36	Country of Manufacturer: USA, EU AND JAPAN	EU	
SPECIFICATION COMPLIANCE			
	TECHNICAL ELIGIBILITY OF THE PRODUCT		Yes
	TECHNICAL ELIGIBILITY OF THE FIRM		Yes
	Bid Status		Responsive

[Handwritten signature]

William Z. August

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THQ HOSPITAL BUREWALA DISTRICT VEHARI**

**TECHNICAL EVALUATION FOR THE PURCHASE OF X Ray Illuminator (Single)
(Qty-12) Estimated Cost. Rs. 2,500/- Each**

Sr# 30B - X Ray Illuminator (Single)

PART-I

Knock Down Criteria (Bid Evaluation)

Sr#	Evaluation Parameters	M/s Noorani Surgical (Pvt) Ltd	M/s Orient Medical
1	Complete Package / tender	Yes	Yes
2	Original Receipt of Tender	Yes	Yes
3	Affidavit from Bidder	Yes	Yes
4	Bid Security	Yes	Yes
5	Bid Validity	Yes	Yes
6	Delivery Period	Yes	Yes
Remarks			
Eligible			

PART-II

Knock Down Criteria (Vendor Evaluation)

1	Valid Authorization Certificate from Manufacturer	Yes	Yes
2	Technical & Engineering Capability	Yes	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes	Yes
4	Vendor Past Performance of Quoted Equipment	Yes	Yes
5	Availability of relevant tools and testing / calibration equipment	Yes	Yes
6	Compliance of Warranty as per tender	Yes	Yes
Remarks			
Eligible			

PART III

Knock Down Criteria (Product Evaluation)






Specification Compliance / Evaluation Parameters	Local	Local
1 Model		

M. Khan
Chairman

L. V. ...

[Signature]

2	Make		Local	
3	Country of Origin		Pakistan	Pakistan
4	Technical Specification			
5	Ultra-Thin X-Ray Film Illuminator using LED Lamps		Yes	Yes
6	Wall mounted		Yes	Yes
7	For one Films of 14"x17" size.		Yes	Yes
8	Thickness: 30 mm or less.		Yes	Yes
9	LED Lamps life: More than 30,000-40,000 hours.		Yes	Yes
10	Easy insertion & removal of the film		Yes	Yes
11	X-Ray film Holder: Rust-free steel clapper (rolling pin) Homogeneous illumination, flicker free.		Yes	Yes
12	Illuminance: 3,500-4,500 cd/m ² .		Yes	Yes
13	On-Off button & separate Brightness control.		Yes	Yes
14	Film sensor to turn-off at some interval after film removal.		Yes	Yes
15	Operating voltage: 220V, 50Hz		Yes	Yes
16	Local Made Subject to Approval of Sample			
		SPECIFICATION COMPLIANCE	N/A	Yes
		TECHNICAL ELIGIBILITY OF THE PRODUCT	NO	yes
		TECHNICAL ELIGIBILITY OF THE FIRM	Yes	yes
		Bid Status	Non-Responsive	Responsive

 Sample not provided
 L-T

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THE HOSPITAL BUREWALA DISTRICT VEHARI**

**TECHNICAL EVALUATION FOR THE PURCHASE OF X Ray Illuminator (Double)
(Qty-08) Estimated Cost. Rs. 4,000/- Each**

Sr# 31B - X Ray Illuminator (Double)

PART-I

Knock Down Criteria (Bid Evaluation)

Sr#	Evaluation Parameters	M/s Noorani Surgical (Pvt) Ltd	M/s Orient Medical
1	Complete Package / tender	Yes	Yes
2	Original Receipt of Tender	Yes	Yes
3	Affidavit from Bidder	Yes	Yes
4	Bid Security	Yes	Yes
5	Bid Validity	Yes	Yes
6	Delivery Period	Yes	Yes
		Remarks	Eligible

PART-II






Knock Down Criteria (Vendor Evaluation)

1	Valid Authorization Certificate from Manufacturer	Yes	Yes
2	Technical & Engineering Capability	Yes	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes	Yes
4	Vendor Past Performance of Quoted Equipment	Yes	Yes
5	Availability of relevant tools and testing / calibration equipment	Yes	Yes
6	Compliance of Warranty as per tender	Yes	Yes
		Remarks	Eligible

PART III

Knock Down Criteria (Product Evaluation)

Specification Compliance / Evaluation Parameters	Local	Local
1 Model	Local	Local

2	Make		Local	Local
3	Country of Origin		Pakistan	Pakistan
4	Technical Specification		Yes	Yes
5	Ultra-Thin X-Ray Film Illuminator using LED Lamps		Yes	Yes
6	Wall mounted		Yes	Yes
7	For two Films of 14"x17" size.		Yes	Yes
8	Thickness: 30 mm or less.		Yes	Yes
9	LED Lamps life: More than 30,000-40,000 hours.		Yes	Yes
10	Easy insertion & removal of the film		Yes	Yes
11	X-Ray film Holder: Rust-free steel clapper (rolling pin) Homogeneous illumination, flicker free.		Yes	Yes
12	Illuminance: 3,500-4,500 cd/m ²		Yes	Yes
13	On-Off button & separate Brightness control.		Yes	Yes
14	Film sensor to turn-off at some interval after film removal.		Yes	Yes
15	Operating voltage: 220V, 50Hz		Yes	Yes
16	Local Made Subject to Approval of Sample			
		SPECIFICATION COMPLIANCE	No	Yes
		TECHNICAL ELIGIBILITY OF THE PRODUCT	No	Yes
		TECHNICAL ELIGIBILITY OF THE FIRM	Yes	Yes
		Bid Status	Non-Responsive	Responsive

Jaam

[Signature]

Chauhan

[Signature]

Sample not provided

[Signature]

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR UPGRADATION OF
THQ HOSPITAL BUREWALA DISTRICT VEHARI
TECHNICAL EVALUATION FOR THE PURCHASE OF SUCTION MACHINE WITH 80L/MIN SUCTION
QTY-10, ESTIMATED COST Rs. 350,000/- EACH**

S# 32B - SUCTION MACHINE WITH 80L/MIN SUCTION

PART-I

S#	Evaluation Parameters	Knock Down Criteria (Bid Evaluation)						
		M/s Eastern Medical	M/s Total Technologies	M/s G-Med	M/s Vial Care	M/s Forcare	M/s Sigma International	M/s Sial Traders
1	Complete Package / tender	Yes	Yes	Yes	Yes	Yes	Yes	Yes
2	Original Receipt of Tender	Yes	Yes	Yes	Yes	Yes	Yes	Yes
3	Affidavit from Bidder	Yes	Yes	Yes	Yes	Yes	Yes	Yes
4	Bid Security	Yes	Yes	Yes	Yes	Yes	Yes	Yes
5	Bid Validity	Yes	Yes	Yes	Yes	Yes	Yes	Yes
6	Delivery Period	Yes	Yes	Yes	Yes	Yes	Yes	Yes
	Remarks	Eligible	Eligible	Eligible	Eligible	Eligible	Eligible	Eligible

PART-II

Knock Down Criteria (Vendor Evaluation)	
1	Valid Authorization Certificate from Manufacturer
2	Technical & Engineering Capability
3	Certificate from Manufacturer about after sale Service through agent or itself
4	Vendor Past Performance of Quoted Equipment
5	Availability of relevant tools and testing / calibration equipment
6	Compliance of Warranty as per tender
	Remarks

PART-III

Knock Down Criteria (Product Evaluation)	
Specification Compliance / Evaluation Parameters	
1	Model
2	Make
3	Country of Origin
	MAXIA SPEED

Victoria Lipos	Eurovac H-50	9.2	Eurovac H-50	MEVACS M-90	New Hospivac 400	Eurovac H-90
Chelron	Hersil	3A	Hersil	Medist Humanne	CAMI	Hersil
Czech Republic	Spain	Italy	Spain	Slovakia	Italy	Spain

[Signature]

Chauhan

[Signature]

4	Certification	CE	CE	CE	CE	CE	CE	CE
5	Technical Specification							
6	Reciprocating, oil free pump mechanism.	Reciprocating Pump	Reciprocating Pump	Reciprocating Pump	Reciprocating Pump	Reciprocating Pump	Yes	Reciprocating Pump
7	Heavy duty Mobile Suction Unit with twin jars (Polysulfone / Polycarbonate type)	Yes	Yes	Yes	Yes	Yes	Yes	Yes
8	Capacity up to 4 or 5 liter each. Autoclave able.	4 liters	4 liters	2 liters	4 liters	2 liters	4	4 liters
9	Aspiration rate up to 80 liters/minutes or more at 650-900mm.	50 LPM @ 5700 mmHg	50 LPM @ 675 mmHg	90 LPM	50 LPM @ 575 mmHg	108 LPM @ 693 mmHg	90 LPM @ 675 mmHg	90 LPM @ 675 mmHg
10	Hg Vacuum continuously adjustable	Yes	Yes	Yes	Yes	Yes	Yes	Yes
11	Triple flow safety device Change over valve	Yes	Yes	Yes	Yes	Yes	Yes	Yes
12	Suction tubing of silicone with coupling connection for each jar	Yes	Yes	Yes	Yes	Yes	Yes	Yes
13	Noise Level 50 dB or less, 220V/50Hz	39.4dB	45dB	45dB	45dB	45dB	46.4dB	50dB
14	Accessories							
15	10 x bacterial filter	Offered	Offered	Offered	Offered	Offered	Offered	Offered
16	Original trolley with lockable wheels.	Offered	Offered	Offered	Offered	Offered	Bullin Trolley	Offered
17	COUNTRY OF ORIGIN: USA, EEC (European Economic Community) OR JAPAN	EEC	EEC	EEC	EEC	EEC	EEC	EEC
	SPECIFICATION COMPLIANCE	N/A	N/A	Yes	N/A	Yes	Yes	Yes
18	TECHNICAL ELIGIBILITY OF THE PRODUCT	No	No	Yes	No	Yes	Yes	Yes
19	TECHNICAL ELIGIBILITY OF THE FIRM	Yes	Yes	Yes	Yes	Yes	Yes	Yes
20	Bid Status	Non-Response	Non-Response	Responsive	Non-Response	Responsive	Responsive	Responsive

[Handwritten signature]

Alan z

Alingalo

[Handwritten signature]

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR UPGRADATION OF
THQ HOSPITAL BUREWALA DISTRICT VEHARI
TECHNICAL EVALUATION FOR THE PURCHASE OF OXYGEN CONCENTRATOR
QTY-1, ESTIMATED COST Rs. 150,000/- EACH**

Sr# 33B - OXYGEN CONCENTRATOR

Knock Down Criteria (Bid Evaluation)

Sr#	Evaluation Parameters	M/s Med Impex
1	Complete Package / tender	Yes
2	Original Receipt of Tender	Yes
3	Affidavit from Bidder	Yes
4	Bid Security	Yes
5	Bid Validity	Yes
6	Delivery Period	Yes
		Remarks
		Eligible

PART-II

Knock Down Criteria (Vendor Evaluation)

1	Valid Authorization Certificate from Manufacturer	Yes
2	Technical & Engineering Capability	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes
4	Vendor Past Performance of Quoted Equipment	Yes
5	Availability of relevant tools and testing / calibration equipment	Yes
6	Compliance of Warranty as per tender	Yes
		Remarks
		Eligible

PART III

Knock Down Criteria (Product Evaluation)

	Specification Compliance / Evaluation Parameters	
1	Model	525KS-LT







2	Make		Drive Devilbiss
3	Country of Origin		Germany
4	Certification		TUV
5	Technical Specification		?? Yes
6	Oxygen concentrator with Built-in oxygen purity indicator.		Yes
7	High and low oxygen alarms.		Yes
8	Metal cannula connection.		Yes
9	Option to connect all types of Humidifier Bottles.		93%
10	Oxygen concentration should be above 90% at all flow settings.		0.5-5 LPM
11	Flow meter scale should be from 0.5 to 5.0 liter per minute.		0.5-5 LPM Yes
12	Outlet pressure should be more than 5.0 PSI.		48dBa Yes
13	Sound level should not be more than 45 dBA.		
14	Country of Manufacturer: USA, EU AND JAPAN		
		SPECIFICATION COMPLIANCE	Yes
		TECHNICAL ELIGIBILITY OF THE PRODUCT	Yes
		TECHNICAL ELIGIBILITY OF THE FIRM	Yes
		Bid Status	Responsive




 Mr. Khan z
 Ayala


 L-K-K-1

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THQ HOSPITAL BUREWALA DISTRICT VEHARI**

**TECHNICAL EVALUATION FROM THE PURCHASE OF CARDIAC RESUSCITATION TROLLEY WITH OXYGEN CYLINDER, SUCKER MACHINE &
DEFIBRILLATOR QTY-04, ESTIMATED COST Rs. 600,000/- EACH**

SMT34B - CARDIAC RESUSCITATION TROLLEY WITH OXYGEN CYLINDER-SUCKER MACHINE AND DEFIBRILLATOR

PART-I

Knock Down Criteria (Bid Evaluation)

S#	Evaluation Parameters	M/s Sigma International	M/s G-Med
1	Complete Package / tender	Yes	Yes
2	Original Receipt of Tender	Yes	Yes
3	Affidavit from Bidder	Yes	Yes
4	Bid Security	Yes	Yes
5	Bid Validity	Yes	Yes
6	Delivery Period	Yes	Yes
	Remarks	Eligible	Eligible

PART-II

Knock Down Criteria (Vendor Evaluation)

1	Valid Authorization Certificate from Manufacturer	Yes
2	Technical & Engineering Capability	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes
4	Vendor Past Performance of Quoted Equipment	Yes
5	Avalability of relevent tools and testing / calibration equipment	Yes
6	Compliance of Warranty as per tender	Yes
	Remarks	Eligible

PART-III

Knock Down Criteria (Product Evaluation)

[Handwritten signatures and notes in the bottom section]

Specification Compliance / Evaluation Parameters			
1	Model	Persolite Inox PCLM	Alfa 01A
2	Make	Francehospital	CFS
3	Country of Origin	Italy	Italy / France
4	Certification		CE
5	Technical Specification		
6	Dished stainless steel / Polymer top.	Yes	Polymer Top Power Coated Steel RAL 7035
7	Approx. 630 x 445mm x 25mm Dished.	?	520mm x 400mm x 60mm
8	Lift-up laminated work flap Approx. 300 x 450	Yes	300X400
9	Three drawers 2 shallow and 1 deep.	Yes	Yes
10	1 Shallow drawer fitted with drawer Tidy Unit.	Yes	Yes
11	Lower cupboard with central locking/securing all drawers and cupboard.	Yes	Yes
12	Operated by cupboard door and use of security seals.	Yes	Yes
13	Double hook stainless steel I.V. Pole.	Yes	Yes
14	Double push handle.	Yes	Yes
15	Defibrillator tray with Biphasic Defibrillator		
16	Model	Reanibex 700	Defigard 400
17	Make	Osatu	Schiller
18	Country of Origin	Spain	Switzerland
19	Low level plastic bumper bar		Yes
20	Quality cushion castors (2 x braking)	Yes	Yes
21	2 x stainless steel cylinder holders for D or E size cylinders.	Yes	Yes
22	Cardiac board 600 x 400 x 55mm with stainless steel housing brackets at rear of trolley.	Yes	Yes
23	Universal rail system fitted to width of trolley.	Yes	Yes
24	6" aneroid sphygmomanometer with adult velcro cuff and rail clamp.	Offered	Yes

[Handwritten signature]

Chair:

[Handwritten signature]

[Handwritten signature]

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE
FOR UPGRADATION OF THQ HOSPITAL BUREWALA DISTRICT VEHARI
TECHNICAL EVALUATION FOR THE PURCHASE OF FILM PROCESSOR
QTY-1, ESTIMATED COST Rs. 975,000/- EACH**

S# 35B - FILM PROCESSOR

PART-I

Knock Down Criteria (Bid Evaluation)

S#	Evaluation Parameters	M/s G-Med	M/s Sial Traders
1	Complete Package / tender	Yes	Yes
2	Original Receipt of Tender	Yes	Yes
3	Affidavit from Bidder	Yes	Yes
4	Bid Security	Yes	Yes
5	Bid Validity	Yes	Yes
6	Delivery Period	Yes	Yes
	Remarks	Eligible	Eligible

PART-II

Knock Down Criteria (Vendor Evaluation)

1	Valid Authorization Certificate from Manufacturer	Yes	Yes
2	Technical & Engineering Capability	Yes	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes	Yes
4	Vendor Past Performance of Quoted Equipment	Yes	Yes
5	Availability of relevant tools and testing / calibration equipment	Yes	Yes
6	Compliance of Warranty as per tender	Yes	Yes
	Remarks	Eligible	Eligible

PART III

Knock Down Criteria (Product Evaluation)

Specification Compliance / Evaluation Parameters

	Remarks	Eligible	Eligible
--	---------	----------	----------

1	Model	XP-9000	Mediphot 900E
2	Make	Daiko Mitech	Colenta Labortechnik
3	Country of Origin	Japan MHLW	Austria CE
4	Certification		
5	Technical Specification		
6	High throughput free Standing type	71 Films/hr	71 Films/hr
7	Continuous roller transport	Yes	Yes
8	Sheet film, 10x10cm-35x43cm sizes (14"x17")	Yes	Yes
9	Medium through around 50 sheets (35"x43")	Yes	Yes
10	Forced hot air circulation method	Yes	Yes
11	Pipng installed in the DEV and FIX racks.	Yes	Yes
12	Processing solution temperature;	Yes	Yes
13	Controlled by the temperature control tank with heater	Yes	Yes
14	heating and wash water cooling;	Yes	Yes
15	Drying temperature; Controlled automatically according to a fixed temperature setting.	Yes	Yes
16	220 V, 50 Hz	Yes	Yes
17	COUNTRY OF ORIGIN, USA EEC OR JAPAN	Japan	EEC
SPECIFICATION COMPLIANCE			
TECHNICAL ELIGIBILITY OF THE PRODUCT		Yes	Yes
TECHNICAL ELIGIBILITY OF THE FIRM		Yes	Yes
Bid Status		Responsive	Responsive

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THQ HOSPITAL BUREWALA DISTRICT VEHARI**

**TECHNICAL EVALUATION FOR THE PURCHASE OF Air Condition Split 1-1/2 Ton
(Qty-60) Estimated Cost. Rs. 60,000/- Each**

Sr# 36B - Air Condition Split 1-1/2 Ton

**PART-I
Knock Down Criteria (Bid Evaluation)**

Sr#	Evaluation Parameters	Remarks
1	Complete Package / tender	No Firm Participated
2	Original Receipt of Tender	
3	Affidavit from Bidder	
4	Bid Security	
5	Bid Validity	
6	Delivery Period	

**PART-II
Knock Down Criteria (Vendor Evaluation)**

1	Valid Authorization Certificate from Manufacturer	
2	Technical & Engineering Capability	
3	Certificate from Manufacturer about after sale Service through agent or itself	
4	Vendor Past Performance of Quoted Equipment	
5	Availability of relevant tools and testing / calibration equipment	
6	Compliance of Warranty as per tender	

**PART III
Knock Down Criteria (Product Evaluation)**

Specification Compliance / Evaluation Parameters		Remarks
1	Model	
2	Make	






1-1-2024

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THQ HOSPITAL BUREWALA DISTRICT VEHARI**

**TECHNICAL EVALUATION FOR THE PURCHASE OF Wooden Bench-02 Seater
(Qty-90) Estimated Cost. Rs. 4,000/- Each**

Sr# 37B - Wooden Bench-02 Seater

PART-I

Knock Down Criteria (Bid Evaluation)

Sr#	Evaluation Parameters	M/s Pieces Engineering & GOS	M/s Noorani Surgical (Pvt) Ltd
1	Complete Package / tender	Yes	Yes
2	Original Receipt of Tender	Yes	Yes
3	Affidavit from Bidder	Yes	Yes
4	Bid Security	No	Yes
5	Bid Validity	Yes	Yes
6	Delivery Period	Yes	Yes
		Remarks	Not-Eligible
			Eligible

PART-II

Knock Down Criteria (Vendor Evaluation)

1	Valid Authorization Certificate from Manufacturer	Yes	Yes
2	Technical & Engineering Capability	Yes	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes	Yes
4	Vendor Past Performance of Quoted Equipment	Yes	Yes
5	Availability of relevant tools and testing / calibration equipment	Yes	Yes
6	Compliance of Warranty as per tender	Yes	Yes
		Remarks	Eligible
			Eligible

PART III

Knock Down Criteria (Product Evaluation)

Specification Compliance / Evaluation Parameters	Local	Local
1 Model	Yes	Yes

2	Make		Local	
3	Country of Origin		Pakistan	Local
4	Technical Specification			
5	MS Pipe/ tube base structure (1"x1½") with aluminum paint.		Yes	Yes
6	Full Wooden Top of rose wood ¾" thick (best quality)		Yes	Yes
7	Size of the bench is 60" x18"x15" (L x H x W) approx.		Yes	Yes
8	Colour :		Yes	Yes
9	The theme colour of all the items is blue, blue shades.			Yes
10	Note: The firm will provide the best brand Vernier Caliper, Screw Gauge, Digital Paint Thickness Meter, Circular Sheet Gauge, Measurement Tape (02 Nos. each) with complete supply.			
11	Local Made Subject to Approval of Sample			
		SPECIFICATION COMPLIANCE	N/A	N/A
		TECHNICAL ELIGIBILITY OF THE PRODUCT	N/A	No
		TECHNICAL ELIGIBILITY OF THE FIRM	Yes	Yes

Handwritten signature

Handwritten signature

Handwritten signature

Handwritten signature

non-Responsive

non-Responsive

Sample is not provided.

Sample is not provided.

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THQ HOSPITAL BUREWALA DISTRICT VEHARI
TECHNICAL EVALUATION FOR THE PURCHASE OF Revolving Office Chair
(Qty-10) Estimated Cost. Rs. 10,000/- Each**

Sr# 38B - Revolving Office Chair

PART-I

Knock Down Criteria (Bid Evaluation)

Sr#	Evaluation Parameters	M/s Noorani Surgical (Pvt) Ltd
1	Complete Package / tender	Yes
2	Original Receipt of Tender	Yes
3	Affidavit from Bidder	Yes
4	Bid Security	Yes
5	Bid Validity	Yes
6	Delivery Period	Yes
Remarks		
Eligible		

PART-II

Knock Down Criteria (Vendor Evaluation)

1	Valid Authorization Certificate from Manufacturer	Yes
2	Technical & Engineering Capability	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes
4	Vendor Past Performance of Quoted Equipment	Yes
5	Availability of relevant tools and testing / calibration equipment	Yes
6	Compliance of Warranty as per tender	Yes
Remarks		
Eligible		

PART III

Knock Down Criteria (Product Evaluation)

Specification Compliance / Evaluation Parameters	Local
1 Model	

[Handwritten signatures and initials]

[Signature] *[Signature]* *[Signature]*

[Signature] *[Signature]*

2	Make	Local
3	Country of Origin	Pakistan
4	Technical Specification	
5	Hydraulic	Yes
6	5 wheels	Yes
7	Adjustable height	Yes
8	Structure made of solid seasoned Shisham wood, finished with NC lacquer seat / back cushioned best quality foam and tapestry cloth.	Yes
9	Local Made Subject to Approval of Sample	
	SPECIFICATION COMPLIANCE	N/C
	TECHNICAL ELIGIBILITY OF THE PRODUCT	N/C
	TECHNICAL ELIGIBILITY OF THE FIRM	Yes
	Bid Status	Non-responsive

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THQ HOSPITAL BUREWALA DISTRICT VEHARI**

**TECHNICAL EVALUATION FOR THE PURCHASE OF Office Chair
(Qty-100) Estimated Cost. Rs. 1,500/- Each**

Sr# 39B - Office Chair

PART-I

Knock Down Criteria (Bid Evaluation)

Sr#	Evaluation Parameters	M/s Noorani Surgical (Pvt) Ltd
1	Complete Package / tender	Yes
2	Original Receipt of Tender	Yes
3	Affidavit from Bidder	Yes
4	Bid Security	Yes
5	Bid Validity	Yes
6	Delivery Period	Yes
Remarks		Eligible

PART-II

Knock Down Criteria (Vendor Evaluation)

1	Valid Authorization Certificate from Manufacturer	Yes
2	Technical & Engineering Capability	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes
4	Vendor Past Performance of Quoted Equipment	Yes
5	Availability of relevant tools and testing / calibration equipment	Yes
6	Compliance of Warranty as per tender	Yes
Remarks		Eligible

PART III

Knock Down Criteria (Product Evaluation)

Specification Compliance / Evaluation Parameters	Model	Local
1		

2	Make		Local
3	Country of Origin		Pakistan
4	Technical Specification		Yes
5	Made of solid wood & upholstered with best quality foam & leather/fabric		Yes
6	M.S pipe powder coated		Yes
7	Polish finish		Yes
8	Size: (600 mm) W x (600 mm) D x (900 mm) H (Approximately)		
9	Local Made Subject to Approval of Sample		
		SPECIFICATION COMPLIANCE	
		TECHNICAL ELIGIBILITY OF THE PRODUCT	NO
		TECHNICAL ELIGIBILITY OF THE FIRM	NO
		Bid Status	Yes

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

Non-Responsive sample not provided

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THQ HOSPITAL BUREWALA DISTRICT VEHARI**

**TECHNICAL EVALUATION FOR THE PURCHASE OF Office Rack
(Qty-20) Estimated Cost. Rs. 2,000/- Each**

Sr# 40B - Office Rack

PART-I

Knock Down Criteria (Bid Evaluation)

Sr#	Evaluation Parameters	M/s Noorani Surgical (Pvt) Ltd
1	Complete Package / tender	Yes
2	Original Receipt of Tender	Yes
3	Affidavit from Bidder	Yes
4	Bid Security	Yes
5	Bid Validity	Yes
6	Delivery Period	Eligible
Remarks		

PART-II

Knock Down Criteria (Vendor Evaluation)

1	Valid Authorization Certificate from Manufacturer	Yes
2	Technical & Engineering Capability	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes
4	Vendor Past Performance of Quoted Equipment	Yes
5	Availability of relevant tools and testing / calibration equipment	Yes
6	Compliance of Warranty as per tender	Eligible
Remarks		

PART III

Knock Down Criteria (Product Evaluation)

Specification Compliance / Evaluation Parameters	Local
1 Model	

2	Make		Local
3	Country of Origin		Pakistan
4	Technical Specification		
5	Structure made of veneer pressed on particle board + solid wood + leatherite wrapped on top		Yes
6	Polish finish		Yes
7	Size: (1530 mm)W x (460 mm)D x (760 mm)H approximately		Yes
8	Local Made Subject to Approval of Sample		
		SPECIFICATION COMPLIANCE	No
		TECHNICAL ELIGIBILITY OF THE PRODUCT	No
		TECHNICAL ELIGIBILITY OF THE FIRM	Yes
		Bid Status	Non-Responsive Sample is not provided

Imran
Amyas

Abbas
Abbas

L-k alden

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THQ HOSPITAL BUREWALA DISTRICT VEHARI**
TECHNICAL EVALUATION FOR THE PURCHASE OF Chair Poshish
(Qty-30) Estimated Cost. Rs. 4,500/- Each

Sr# 41B - Chair Poshish

PART-I

Knock Down Criteria (Bid Evaluation)

Sr#	Evaluation Parameters	M/s Noorani Surgical (Pvt) Ltd
1	Complete Package / tender	Yes
2	Original Receipt of Tender	Yes
3	Affidavit from Bidder	Yes
4	Bid Security	Yes
5	Bid Validity	Yes
6	Delivery Period	Yes
PART-II		
Knock Down Criteria (Vendor Evaluation)		
1	Valid Authorization Certificate from Mnaufacturer	Yes
2	Technical & Engineering Capability	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes
4	Vendor Past Performance of Quoted Equipment	Yes
5	Availability of relevent tools and testing / calibration equipment	Yes
6	Compliance of Warrantly as per tender	Yes
PART III		
Knock Down Criteria (Product Evaluation)		
Specification Compliance / Evaluation Parameters		
1	Model	Local

Remarks

Eligible

Remarks

Eligible

2	Make		Local
3	Country of Origin		Pakistan
4	Technical Specification		
5	Chair Poshish		Yes
6	Local Made Subject to Approval of Sample		
		SPECIFICATION COMPLIANCE	
		TECHNICAL ELIGIBILITY OF THE PRODUCT	No
		TECHNICAL ELIGIBILITY OF THE FIRM	No
		Bid Status	Yes

*Non-Responsive
Sample not provided*

Signature
Signature

Signature

Signature
Signature

M-Kabeer

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THQ HOSPITAL BUREWALA DISTRICT VEHARI**

**TECHNICAL EVALUATION FOR THE PURCHASE OF Fiber Glass Chair
(Qty-20 set each of 3 seats) Estimated Cost. Rs. 10,000/- Each set**

SI# 42B - Fiber Glass Chair

PART-I

Knock Down Criteria (Bid Evaluation)

Sr#	Evaluation Parameters	M/s Noorani Surgical (Pvt) Ltd
1	Complete Package / tender	Yes
2	Original Receipt of Tender	Yes
3	Affidavit from Bidder	Yes
4	Bid Security	Yes
5	Bid Validity	Yes
6	Delivery Period	Eligible
PART-II		
Knock Down Criteria (Vendor Evaluation)		
1	Valid Authorization Certificate from Manufacturer	Yes
2	Technical & Engineering Capability	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes
4	Vendor Past Performance of Quoted Equipment	Yes
5	Availability of relevant tools and testing / calibration equipment	Yes
6	Compliance of Warranty as per tender	Yes
PART III		
Knock Down Criteria (Product Evaluation)		
1	Model	Local







2	Make		Local
3	Country of Origin		Pakistan
4	Technical Specification		
5	Local/Imported Subject to Approval of Sample		
		SPECIFICATION COMPLIANCE	
		TECHNICAL ELIGIBILITY OF THE PRODUCT	N/A
		TECHNICAL ELIGIBILITY OF THE FIRM	N/A
		Bid Status	Yes

*Non-Responsive.
Sample not provided*

Handwritten signatures and initials:
 - Top left: *Handwritten signature*
 - Middle left: *Handwritten signature*
 - Middle right: *Handwritten signature*
 - Bottom right: *Handwritten signature*
 - Far right: *Handwritten signature*

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THE HOSPITAL BUREWALA DISTRICT VEHARI**

**TECHNICAL EVALUATION FOR THE PURCHASE OF Office Table
(Qty-20) Estimated Cost. Rs. 10,000/- Each**

Sr# 43B - Office Table

PART-I

Knock Down Criteria (Bid Evaluation)

Sr#	Evaluation Parameters	M/s Noorani Surgical (Pvt) Ltd
1	Complete Package / tender	Yes
2	Original Receipt of Tender	Yes
3	Affidavit from Bidder	Yes
4	Bid Security	Yes
5	Bid Validity	Yes
6	Delivery Period	Eligible
Remarks		
PART-II		
Knock Down Criteria (Vendor Evaluation)		

1	Valid Authorization Certificate from Manufacturer	Yes
2	Technical & Engineering Capability	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes
4	Vendor Past Performance of Quoted Equipment	Yes
5	Availability of relevant tools and testing / calibration equipment	Yes
6	Compliance of Warranty as per tender	Yes
Remarks		
Eligible		

PART III

Knock Down Criteria (Product Evaluation)

Specification Compliance / Evaluation Parameters	Model	Local
1		

Muhammad Asif

[Signature]

[Signature]

[Signature]

2	Make	Local
3	Country of Origin	Pakistan
4	Technical Specification	
5	Desk plus side rack and drawer pedestal	Yes
6	Structure made of veneer pressed on particle board both sides	Yes
7	Legs with M.S pipe powder coated	Yes
8	Leathrie wrapped on top	Yes
9	Polish finish	Yes
10	Desk: W(2400 mm) *D(1050 mm) *H(760 mm) Approximately	Yes
11	Rack (1500 mm) *D(475 mm) *H(760 mm) Approximately	Yes
12	Drawer Pedestal: W(400 mm) *D(460 mm) *H(650 mm) Approximately	Yes
13	Local Made Subject to Approval of Sample	
	SPECIFICATION COMPLIANCE	No
	TECHNICAL ELIGIBILITY OF THE PRODUCT	No
	TECHNICAL ELIGIBILITY OF THE FIRM	Yes
	Bid Status	Non-Responsive

Non-Responsive
Sample is not provided







**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THQ HOSPITAL BUREWALA DISTRICT VEHARI**

**TECHNICAL EVALUATION FOR THE PURCHASE OF Examination Couch
(Qty-20) Estimated Cost. Rs. 8,000/- Each**

Sr# 44B - Examination Couch

PART-I

Knock Down Criteria (Bid Evaluation)

Sr#	Evaluation Parameters	M/s Thankful Agencies	M/s Noorani Surgical (Pvt) Ltd	M/s Orient Medical	M/s Pieces Engineering & GOS
1	Complete Package / tender	Yes	Yes	Yes	Yes
2	Original Receipt of Tender	Yes	Yes	Yes	Yes
3	Affidavit from Bidder	Yes	Yes	Yes	No
4	Bid Security	Yes	Yes	Yes	Yes
5	Bid Validity	Yes	Yes	Yes	Yes
6	Delivery Period	Yes	Yes	Yes	Yes
	Remarks	Eligible	Eligible	Eligible	Not-Eligible

PART-II

Knock Down Criteria (Vendor Evaluation)

1	Valid Authorization Certificate from Manufacturer	Yes	Yes	Yes	Yes
2	Technical & Engineering Capability	Yes	Yes	Yes	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes	Yes	Yes	Yes
4	Vendor Past Performance of Quoted Equipment	Yes	Yes	Yes	Yes
5	Availability of relevant tools and testing / calibration equipment	Yes	Yes	Yes	Yes
6	Compliance of Warranty as per tender	Yes	Yes	Yes	Yes
	Remarks	Eligible	Eligible	Eligible	Eligible

PART III

Knock Down Criteria (Product Evaluation)

Specification Compliance / Evaluation Parameters				
1	Model	Local		Local
2	Make	Thankful Agencies	Local	Local
3	Country of Origin	Pakistan	Pakistan	Pakistan
4	Technical Specification			
5	Desk plus side rack and drawer pedestal	Yes	Yes	Yes
6	Structure made of veneer pressed on particle board both sides	Yes	Yes	Yes
7	Legs with M.S pipe powder coated	Yes	Yes	Yes
8	Leathrite wrapped on top	Yes	Yes	Yes
9	Polish finish	Yes	Yes	Yes
10	Desk: W(2400 mm) *D(1050 mm) *H(760 mm) Approximately	Yes	Yes	Yes
11	Rack (1500 mm) *D(475 mm) *H(760 mm) Approximately	Yes	Yes	Yes
12	Drawer Pedestal: W(400 mm) *D(460 mm) *H(650 mm) Approximately	Yes	Yes	Yes
13	Local Made Subject to Approval of Sample			
SPECIFICATION COMPLIANCE				
TECHNICAL ELIGIBILITY OF THE PRODUCT				
TECHNICAL ELIGIBILITY OF THE FIRM				
Bid Status				
		Yes	No	No
		Yes	No	No
		Yes	Yes	Yes

Handwritten signature

Handwritten signature

Handwritten signature

Sample not provided

Low quality of sample.

Sample is not provided

Handwritten signature

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THQ HOSPITAL BUREWALA DISTRICT VEHARI
TECHNICAL EVALUATION FOR THE PURCHASE OF Bed Sheet
(Qty-500) Estimated Cost. Rs. 550/- Each**

Sr # 46B - Bed Sheet






PART-I

Knock Down Criteria (Bid Evaluation)

Sr#	Evaluation Parameters	M/s Pieces Engineering & GOS	M/s Noorani Surgical (Pvt) Ltd
1	Complete Package / tender	Yes	Yes
2	Original Receipt of Tender	Yes	Yes
3	Affidavit from Bidder	No	Yes
4	Bid Security	Yes	Yes
5	Bid Validity	Yes	Yes
6	Delivery Period	Not-Eligible	Eligible
PART-II			
Knock Down Criteria (Vendor Evaluation)			

1	Valid Authorization Certificate from Manufacturer	Yes	Yes
2	Technical & Engineering Capability	Yes	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes	Yes
4	Vendor Past Performance of Quoted Equipment	Yes	Yes
5	Availability of relevant tools and testing / calibration equipment	Yes	Yes
6	Compliance of Warranty as per tender	Eligible	Eligible
PART III			
Knock Down Criteria (Product Evaluation)			

Specification Compliance / Evaluation Parameters	Local	Local
1 Model	Local	Local

2	Make		Local	
3	Country of Origin		Pakistan	Pakistan
4	Technical Specification			
5	Bed Sheet Length: 108" Bed Sheet Width: 66"		Yes	Yes
6	With Govt. of the Punjab insignia (one in center) and 1.5" sides woven.		Yes	Yes
7	Bed sheets should be printed with insignia of Govt. Of the Punjab name and Hospital.		Yes	Yes
8	Colour: White		Yes	Yes
9	Best Quality Subject to Approval of Sample			
		SPECIFICATION COMPLIANCE		
		TECHNICAL ELIGIBILITY OF THE PRODUCT	N/A	N/A
		TECHNICAL ELIGIBILITY OF THE FIRM	N/A	N/A
		Bid Status	Non-Responsive	Non-Responsive

Handwritten signature

Handwritten signature

Handwritten signature

Handwritten signature

Handwritten signature

Sample is not provided

Sample is not provided

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THQ HOSPITAL BUREWALA DISTRICT VEHARI**

**TECHNICAL EVALUATION FOR THE PURCHASE OF Blanket
(Qty-90) Estimated Cost. Rs. 4,000/- Each**

Sr # 47B - Blanket

PART-I

Knock Down Criteria (Bid Evaluation)

Sr#	Evaluation Parameters	M/s Med Impex	M/s Noorani Surgical (Pvt) Ltd	M/s Pieces Engineering & GOS
1	Complete Package / tender	Yes	Yes	Yes
2	Original Receipt of Tender	Yes	Yes	Yes
3	Affidavit from Bidder	Yes	Yes	Yes
4	Bid Security	Yes	Yes	No
5	Bid Validity	Yes	Yes	Yes
6	Delivery Period	Yes	Yes	Yes
	Remarks	Eligible	Eligible	Not-Eligible

PART-II

Knock Down Criteria (Vendor Evaluation)

1	Valid Authorization Certificate from Manufacturer	Yes	Yes	Yes
2	Technical & Engineering Capability	Yes	Yes	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes	Yes	Yes
4	Vendor Past Performance of Quoted Equipment	Yes	Yes	Yes
5	Availability of relevant tools and testing / calibration equipment	Yes	Yes	Yes
6	Compliance of Warranty as per tender	Yes	Yes	Yes
	Remarks	Eligible	Eligible	Eligible

PART III

Knock Down Criteria (Product Evaluation)

Specification Compliance / Evaluation Parameters	Local	Local	Local
1 Model			

Chair

Local

2	Make	Local	Local	Local
3	Country of Origin	Pakistan	Pakistan	Pakistan
4	Technical Specification			
5	Fine quality single bed double ply blankets with cover	Yes	Yes	Yes
6	Colour: Red color/Blue/As per requirement	Yes	Yes	Yes
7	The supplier shall also provide the three (03) stitched covers for each blanket.	Yes	Yes	Yes
8	Best Quality Subject to Approval of Sample			
		SPECIFICATION COMPLIANCE		
		TECHNICAL ELIGIBILITY OF THE PRODUCT		
		TECHNICAL ELIGIBILITY OF THE FIRM		
		Bid Status		

Handwritten signature

Handwritten signature

Handwritten signature

Handwritten signature

Handwritten signature

Non-Responsive sample is not provided

Non-Responsive Bid security is required.

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THQ HOSPITAL BUREWALA DISTRICT VEHARI**

**TECHNICAL EVALUATION FOR THE PURCHASE OF Pillow with Pillow Cover
(Qty-90) Estimated Cost. Rs. 500/- Each**

Sr # 488 - Pillow with Pillow Cover

PART-I

Knock Down Criteria (Bid Evaluation)

Sr#	Evaluation Parameters	M/s Pieces Engineering & GOS	M/s Noorani Surgical (Pvt) Ltd
1	Complete Package / tender	Yes	Yes
2	Original Receipt of Tender	Yes	Yes
3	Affidavit from Bidder	Yes	Yes
4	Bid Security	Yes	Yes
5	Bid Validity	Yes	Yes
6	Delivery Period	Eligible	Eligible
Remarks			
PART-II			
Knock Down Criteria (Vendor Evaluation)			

1	Valid Authorization Certificate from Manufacturer	Yes	Yes
2	Technical & Engineering Capability	Yes	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes	Yes
4	Vendor Past Performance of Quoted Equipment	Yes	Yes
5	Availability of relevant tools and testing / calibration equipment	Yes	Yes
6	Compliance of Warranty as per tender	Yes	Yes
Remarks			
Eligible			

PART III

Knock Down Criteria (Product Evaluation)

Specification Compliance / Evaluation Parameters	Local	Local
1 Model	Local	Local

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THQ HOSPITAL BUREWALA DISTRICT VEHARI
TECHNICAL EVALUATION FOR THE PURCHASE OF Mattress with Cover
(Qty-90) Estimated Cost. Rs. 10,000/- Each**

Sr # 49B - Mattress with Cover

PART-I

Knock Down Criteria (Bid Evaluation)

Sr#	Evaluation Parameters	M/s Thankful Agencies	M/s Noorani Surgical (Pvt) Ltd	M/s Pieces Engineering & GOS
1	Complete Package / tender	Yes	Yes	Yes
2	Original Receipt of Tender	Yes	Yes	Yes
3	Attidavit from Bidder	Yes	Yes	No
4	Bid Security	Yes	Yes	Yes
5	Bid Validity	Yes	Yes	Yes
6	Delivery Period	Yes	Yes	Yes
Remarks				
		Eligible	Eligible	Not-Eligible

PART-II

Knock Down Criteria (Vendor Evaluation)






1	Valid Authorization Certificate from Manufacturer	Yes	Yes	Yes
2	Technical & Engineering Capability	Yes	Yes	Yes
3	Certificate from Manufacturer about after sale Service through agent or itself	Yes	Yes	Yes
4	Vendor Past Performance of Quoted Equipment	Yes	Yes	Yes
5	Availability of relevent tools and testing / calibration equipment	Yes	Yes	Yes
6	Compliance of Warranty as per tender	Yes	Yes	Yes
Remarks				
		Eligible	Eligible	Eligible

PART III

Knock Down Criteria (Product Evaluation)

Specification Compliance / Evaluation Parameters

1	Model	Local	Local	Local
2	Make	Thankful Agencies Pakistan	Local Pakistan	Local Pakistan
3	Country of Origin			
4	Technical Specification			
5	Best quality (1st Quality) mattress foam of 4" thickness having density of 35-36 kg/m ³ .	Yes	Yes	Yes
6	Water proof Reverse cover having best quality jersey and Zipper	Yes	Yes	Yes
7	Best Quality Subject to Approval of Sample			
		SPECIFICATION COMPLIANCE		
		TECHNICAL ELIGIBILITY OF THE PRODUCT		
		TECHNICAL ELIGIBILITY OF THE FIRM		
		Bid Status	Responsive	Non-Responsive

Non-Responsive
 Sample not provided
 Non-Responsive
 Sample not provided