TECHNICAL EVALUATION FOR THE PURCHASE OF BIOMEDICAL EQUIPMENT FOR THQ HOSPITAL BUREWAL

	HI NOT INTEREST OF THE PROPERTY OF THE PROPERT	A WIENT FOR THE HOSPITAL BUREWALA	
Packao	BID EVALUATION SHEET		
Tarvas	acrage Ny render Number:		
Name of The	of The Equipment & Quantity: ACT (Activated Clotting Machine) Qty-01, Estimated Cost. Rs. 250,000/-		
Tarry			
KONOC	KNOCK DOWN CRITERIA (COMMECIAL EVALUATION)		
All Eva	All Evaluation Parametrs defined below are mandatory for Complaince		
Sr No.	Evaluation Parameters		
	Legal registered entity	M/s Sami Enterprises	M/s Ali Traders
2	Active Tax Payer (NTN & GST) as per clause 3 TTR	Yes	Yes
w		Yes	Yes
4	Manufacturer/Sole Distributor	Yes	Yes
01	Past performance	No	No
	Rainer Process and the second	No	No
J	The care of the care ment of Annual Lumover	Yes	Yes
1	Complete Item with accessories and optional	Yes	Yes
00	Certificate from the Manufacturer regarding after Sales Services and Installation	No	3
9	Technical & Engineering Capability of firm	No	No
10	Testing/Calibration equipment	Zo	No.
11	Affidavit from Bidder (as per Supplier Declaration Template)	- 10	V
12	Bid Validity	V	Y IS
13	Delivery Period	***	TCS
14	Compliance of Warranty as per tender	Yes	Yes
	REMARKS	Not-Eligible for further evaluation of PART-II	Not-Eligible for further evaluation of PART-II
Part II			
			X

Knoc	k Down Criteria (Product Evaluation)		
All Ev	aluation Parametrs defined below are mandatory for Compliance		
SrNo	Specification Compliance / Evaluation Parameters		
2	ACT Machine		
	Company Name	M/s Sami Enterprises	M/s Ali Trader
	Brand		
	Model		
	Country of Manufacturer		
	Country of origin of product/ Model Number		
	Product Local Market Business History as mentioned in criteria		
	Compliance with defined quality standards (FDA 510k, MDD, Jp MHLW as mentioned in the specifications)		
Sr#	Specifications/Sample Evaluation:		
	TECHNICAL SPECIFICATIONS:		
1	Microprocessor control including tests		
2	PT		
3	APTT		
4	ACT		
5	Or any other Used on both fresh and citrated whole blood sample		
6	LCD display		
7	Data storage for at-least 300 samples		
8	Quality Control Check		
9	Result printing capability		
10	Accessories:		
11	Complete with at least 300 cartridge or more		
12	Optional (if any):		

13	Dual / Single Chamber		
14	Country of Manufacturer: USA/EU/Japan		
15	Country of Origin; USA/EU/Japan		
16	Quality Certificate: FDA 510k/CE MDD/JpMHLW		
17	Mode: DDP		
18	Warranty: 05 Years		
	TECHNICAL ELIGIBILITY OF THE PRODUCT	Not-Eligible	Not-Eligible
	TECHNICAL ELIGIBILITY OF THE FIRM	Not-Eligible	Not-Eligible
	BID STATUS	Non-Responsive	Non-Responsive